

# Standardized endpoint definitions for transcatheter aortic valve implantation clinical trials: a consensus report from the Valve Academic Research Consortium<sup>†</sup>

**Martin B. Leon\***, **Nicolo Piazza**, **Eugenia Nikolsky**, **Eugene H. Blackstone**, **Donald E. Cutlip**, **Arie Pieter Kappetein**, **Mitchell W. Krucoff**, **Michael Mack**, **Roxana Mehran**, **Craig Miller**, **Marie-angèle Morel**, **John Petersen**, **Jeffrey J. Popma**, **Johanna J.M. Takkenberg**, **Alec Vahanian**, **Gerrit-Anne van Es**, **Pascal Vranckx**, **John G. Webb**, **Stephan Windecker**, and **Patrick W. Serruys**

Columbia University Medical Center, Center for Interventional Vascular Therapy, 173 Fort Washington Avenue, Heart Center, 2nd floor, New York, NY 10032, USA.

Received 8 July 2010; revised 30 September 2010; accepted 6 October 2010

## Objectives

To propose standardized consensus definitions for important clinical endpoints in transcatheter aortic valve implantation (TAVI), investigations in an effort to improve the quality of clinical research and to enable meaningful comparisons between clinical trials. To make these consensus definitions accessible to all stakeholders in TAVI clinical research through a peer reviewed publication, on behalf of the public health.

## Background

Transcatheter aortic valve implantation may provide a worthwhile less invasive treatment in many patients with severe aortic stenosis and since its introduction to the medical community in 2002, there has been an explosive growth in procedures. The integration of TAVI into daily clinical practice should be guided by academic activities, which requires a harmonized and structured process for data collection, interpretation, and reporting during well-conducted clinical trials.

## Methods and results

The Valve Academic Research Consortium established an independent collaboration between Academic Research organizations and specialty societies (cardiology and cardiac surgery) in the USA and Europe. Two meetings, in San Francisco, California (September 2009) and in Amsterdam, the Netherlands (December 2009), including key physician experts, and representatives from the US Food and Drug Administration (FDA) and device manufacturers, were focused on creating consistent endpoint definitions and consensus recommendations for implementation in TAVI clinical research programs. Important considerations in developing endpoint definitions included (i) respect for the historical legacy of surgical valve guidelines; (ii) identification of pathophysiological mechanisms associated with clinical events; (iii) emphasis on clinical relevance. Consensus criteria were developed for the following endpoints: mortality, myocardial infarction, stroke, bleeding, acute kidney injury, vascular complications, and prosthetic valve performance. Composite endpoints for TAVI safety and effectiveness were also recommended.

## Conclusion

Although consensus criteria will invariably include certain arbitrary features, an organized multidisciplinary process to develop specific definitions for TAVI clinical research should provide consistency across studies that can facilitate the

<sup>†</sup>The Valve Academic Research Consortium (VARC) consists of representatives from several independent Academic Research Organizations, several Surgery and Cardiology Societies, members of the U.S. Food and Drug Administration, and several independent experts (Appendices 1 and 2).

\* Corresponding author. Tel: +1 212 342 3617, Fax: +1 212 342 3660, Email: [mleon@crf.org](mailto:mleon@crf.org)

Published on behalf of the European Society of Cardiology. The article has been co-published in the Journal of the American College of Cardiology. All rights reserved. © The Author 2010. For permissions please email: [journals.permissions@oup.com](mailto:journals.permissions@oup.com).

The online version of this article has been published under an open access model. Users are entitled to use, reproduce, disseminate, or display the open access version of this article for non-commercial purposes provided that the original authorship is properly and fully attributed; the Journal, Learned Society and Oxford University Press are attributed as the original place of publication with correct citation details given; if an article is subsequently reproduced or disseminated not in its entirety but only in part or as a derivative work this must be clearly indicated. For commercial re-use, please contact [journals.permissions@oup.com](mailto:journals.permissions@oup.com)

evaluation of this new important catheter-based therapy. The broadly based consensus endpoint definitions described in this document may be useful for regulatory and clinical trial purposes.

## Keywords

Transcatheter aortic valve implantation

## Introduction

Since the introduction of transcatheter aortic valve implantation (TAVI) in 2002<sup>1</sup>, there has been increasing interest in the field of catheter-based treatment of high-surgical-risk patients with symptomatic aortic stenosis (AS).<sup>2–7</sup> Introduction of this new technology should ideally follow the standard bench-to bedside evidence-based medicine pattern, starting with pre-clinical testing and advancing to clinical investigations. Unfortunately, the explosive growth of TAVI (Figure 1) has created a ‘clinical data conundrum’: investigators were not prepared to optimally organize and interpret clinical data for this radically different treatment, rendering thoughtful assessment of clinical trial outcomes difficult and inter-study results comparisons problematic.<sup>8–11</sup>


Surgical valve clinical research guidelines have been developed using a more traditional ‘multi-society approach’, have been revised approximately every 10 years, incorporate not merely clinical endpoints but also issues such as structural valve deterioration and non-structural valve dysfunction, and often divide clinical events into those that are valve and non-valve related.<sup>12</sup> Interventional cardiology has a tradition of agreed upon clinical endpoint definitions and clinical trial methodologies<sup>13,14</sup> and recently has incorporated a consensus process to standardize key endpoint definitions by convening an Academic Research Consortium (ARC) among Academic Research Organizations (AROs) from the USA and Europe joined by representatives from the USFDA and device manufacturers.<sup>14</sup> The ARC process demonstrated the

power of a well-managed international goal-directed academic consortium collaborating effectively with the FDA and industry to establish consensus clinical endpoint definitions and to improve the conduct of clinical research.

In the spirit of the ARC-mission statement,<sup>14</sup> the ‘Valve Academic Research Consortium’ (VARC) was organized as an amalgam of the ARO and multi-society guideline models with strong participation from independent experts, the FDA, and medical device manufacturers (Appendices 1 and 2). Two in-person meetings on 19 September 2009 in San Francisco, CA and on 5–6 December 2009 in Amsterdam, the Netherlands, involving VARC study group members and invited guests (including the FDA and industry representatives) provided much of the substantive discussion from which this consensus manuscript was derived.

The goals of VARC are to combine the expertise of surgeons, interventionalists, medical cardiologists, clinical trialists, and other specialists (representing relevant disciplines including echocardiography, vascular medicine, and neurology) to arrive at a consensus for (i) *selecting appropriate clinical endpoints* reflecting device, procedure and patient-related effectiveness and safety, and (ii) *standardizing definitions for single and composite clinical endpoints*.

Importantly, this first consensus manuscript was not intended as a ‘guidelines statement’ or a ‘guidance document’, but rather should be viewed as a roadmap to facilitate the standardization of future TAVI and other aortic valve clinical research



**Figure 1** Current generation transcatheter aortic valve therapies. (A) balloon-expandable, stainless steel support structure, bovine pericardial valve; (B) self-expanding, nitinol support structure, porcine pericardial valve.

## Principles for selecting and defining clinical endpoints for transcatheter aortic valve implantation investigations: general considerations

### Criteria for endpoint definitions

The definitions of major clinical endpoints must follow a multi-step thought process.

- Each major endpoint should address issues that establish either the *safety* and/or the *effectiveness* of the proposed new therapy.
  - *Safety* is characterized by the avoidance of device-related or procedural complications.
  - *Effectiveness* is a more complex descriptor, as it encompasses both the avoidance of negative disease-related outcomes and objective measures of clinical functional benefit.
- The endpoints should relate short- and long-term *pathophysiological mechanisms* to meaningful clinical events.
- Endpoint definitions must be consistent with the body of published literature, but still reflect unique or evolving aspects of the new therapy.
- The emphasis should be on definitions that accurately represent *essential patient-oriented clinical outcomes*.
- The endpoints must be well defined (preferably through blinded adjudication processes) such that they can be to be subjected to statistical analysis.

It is helpful to reference a standardized definition format regarding (i) the specific treatment, (ii) the place of occurrence, (iii) the time of occurrence, and (iv) the specific type of endpoint.

### Device, procedure, and patient-oriented outcomes

Endpoint definitions for TAVI will in most cases be characterized in relation to the specific implant device, the implant procedure, and the resultant patient-oriented outcomes, which can occur at any time after the procedure. During the early phases of therapy development, particular attention must be directed to the safety and performance of the device. Therefore, VARC tries to strike a compromise by also elucidating device and procedure-related events, which are essential to the understanding of a new class of catheter-based therapies.

Since TAVI is fundamentally the placement of a prosthetic aortic valve and will be compared with surgical aortic valve replacement (AVR), tradition should be respected and crucial endpoints such as all-cause mortality and device durability must be assessed longitudinally for the life of the implant.<sup>12</sup> However, primary clinical endpoints used in pivotal clinical trials for regulatory approval of TAVI devices should incorporate a shorter time domain of 1 to 2 years after the index procedure. These recommended shorter time horizons should not discourage the standard long-term follow-up procedures for prosthetic heart valves.

## Proposed safety and efficacy endpoints

### Mortality

*All-cause mortality* in surgical clinical trials has become the 'gold standard' in previously published consensus and guideline documents.<sup>12</sup> The advantage of reporting all-cause mortality is that it is both objective (without bias) and pragmatic from the standpoint of ascertainment and adjudication. However, the use of all-cause mortality in high-risk TAVI patients may be misleading, resulting in disproportionate reporting of mortal events unrelated to either the treatment device or the procedure. Therefore, VARC proposes to use all-cause mortality as a primary clinical endpoint, but also recommends further subdivision of mortality, specifically denoting *cardiovascular mortality* as an important secondary endpoint (Table 1). Of note, 'unknown' deaths should be considered as cardiovascular in origin and to improve the ascertainment of death, the social security death index or national death registries should be utilized in cases of patients lost to follow-up.

Consistent with surgical guidelines and surgical clinical trial practices,<sup>12</sup> mortality should be formally assessed and reported at 30 days after the index procedure (or longer if the patient was not discharged from the treatment hospital or a secondary convalescent facility). Since there may be either unknown or under-reporting of early device failure modes, a more appropriate duration for all-cause mortality as a primary endpoint in TAVI clinical trials is 1 year after the index procedure. After 1 year, mortality should be recorded at yearly intervals for a minimum of 5 years, or ideally, for the duration of the prosthetic valve implant, in the form of well-defined post-approval surveillance registries.

### Myocardial infarction

In 2007, the joint ESC/ACC/AHA/WHF task force for the redefinition of Myocardial infarction (MI) established diagnostic criteria and updated guidelines for a universal MI definition to be used in clinical trials.<sup>13</sup> This universal MI definition is highly sensitive, relying heavily on the measurement of cardiac biomarkers (preferably troponin). Conversely, surgical valve guidelines have adopted a

**Table 1 Mortality**

#### Cardiovascular mortality

Any one of the following criteria:

Any death due to proximate cardiac cause (e.g. MI, cardiac tamponade, worsening heart failure)

Unwitnessed death and death of unknown cause

All procedure-related deaths, including those related to a complication of the procedure or treatment for a complication of the procedure

Death caused by non-coronary vascular conditions such as cerebrovascular disease, pulmonary embolism, ruptured aortic aneurysm, dissecting aneurysm, or other vascular disease

**Table 2 Myocardial infarction**

|   |
|---|
| <p>Peri-procedural MI (<math>\leq 72</math> h After the index procedure)</p> <p>New ischaemic symptoms (e.g. chest pain or shortness of breath), or new ischaemic signs (e.g. ventricular arrhythmias, new or worsening heart failure, new ST-segment changes, haemodynamic instability, or imaging evidence of new loss of viable myocardium or new wall motion abnormality),</p> <p>AND</p> <p>Elevated cardiac biomarkers (preferably CK-MB) within 72 h after the index procedure, consisting of two or more post-procedure samples that are <math>&gt;6-8</math> h apart with a 20% increase in the second sample and a peak value exceeding 10x the 99th percentile upper reference limit (URL), or a peak value exceeding 5x the 99th percentile URL with new pathological Q waves in at least two contiguous leads.</p> <hr/> <p>Spontaneous MI (<math>&gt;72</math> h after the index procedure)</p> <p>Any one of the following criteria:</p> <ul style="list-style-type: none"> <li>Detection of rise and/or fall of cardiac biomarkers (preferably troponin) with at least one value above the 99th percentile URL, together with evidence of myocardial ischaemia with at least one of the following: <ul style="list-style-type: none"> <li>ECG changes indicative of new ischaemia [new ST-T changes or new left bundle branch block (LBBB)]</li> <li>New pathological Q waves in at least two contiguous leads</li> <li>Imaging evidence of new loss of viable myocardium or new wall motion abnormality</li> </ul> </li> <li>Sudden, unexpected cardiac death, involving cardiac arrest, often with symptoms suggestive of myocardial ischaemia, and accompanied by presumably new ST elevation, or new LBBB, and/or evidence of fresh thrombus by coronary angiography and/or at autopsy, but death occurring before blood samples could be obtained, or at a time before the appearance of cardiac biomarkers in the blood.</li> <li>Pathological findings of an acute myocardial infarction.</li> </ul> |
|---|

'minimalist' approach to MI definitions, usually ignoring biomarker diagnoses and excluding both intra-operative and post-operative MIs, unless the MI was caused by a coronary embolus.<sup>12</sup>

Valve Academic Research Consortium proposes a more 'centrist' approach to MI definitions after TAVI, recognizing that many patients have coexistent aortic valve and coronary artery disease,<sup>15</sup> which requires an MI definition that does not exclude peri-procedural or late MIs that may impact patient outcomes. Valve Academic Research Consortium proposes to define *peri-procedural MI* as an acute ischaemic event that is associated with documented and *clinically significant myocardial necrosis* (Table 2). This definition does not include ischaemic events after TAVI or surgery defined solely by biomarker elevations without a clinically evident ischaemic insult. Since troponin measurements are an extremely sensitive biomarker of myocardial necrosis, VARC recommends that CPK-MB should be the peri-procedural biomarker of choice in TAVI clinical trials. Biomarker samples ideally should be obtained at baseline, twice after the procedure (separated by at least 6 h), and if still elevated, daily thereafter until values are declining. Since TAVI may involve open surgical procedures (e.g. transapical access), the biomarker diagnosis of peri-procedural MI requires a  $>20\%$  increase

in the second post-procedure sample and a threshold elevation of ten times the upper normal range. An electrocardiogram should be collected at baseline and at least once after the procedure prior to discharge to document the presence or absence of new Q-waves. The peri-procedural interval is inclusive of all events that begin within 72 h of the index procedure. Acute ischaemic events occurring after 72 h are considered *spontaneous MIs* and are defined in accordance with the universal MI guidelines,<sup>13</sup> as further modified by ARC<sup>14</sup> (Table 2). Finally, a confirmed coronary embolus, occurring at any time, should be reported as an independent event, if biomarker changes and associated findings fulfil definition criteria.

## Stroke

Recently, two reports have indicated a high frequency of new perfusion abnormalities (presumably embolic) detected by diffusion-weighted magnetic resonance imaging (MRI) studies soon after TAVI,<sup>16,17</sup> although the clinical significance of these early perfusion defects remains unclear. Strokes during and after TAVI may occur due to embolic events from multiple sources, procedure-related aortic dissections, ischaemia from hypotension, or haemorrhagic complications associated with adjunctive pharmacotherapy. Insights on stroke definitions are in a state of evolution and VARC examined viewpoints derived from several sources, including recent multi-society consensus documents<sup>12,18,19</sup> and multi-center randomized trials, in which stroke was an important endpoint.<sup>20-29</sup> Valve Academic Research Consortium considered five important issues in arriving at clinically relevant stroke definitions, as follows: (i) a clinical diagnosis of stroke which ruled out metabolic or toxic encephalopathy, pharmacological influences, and non-central neurological symptoms, (ii) the role of neuroimaging studies for confirmation of the diagnosis, (iii) the distinction of stroke vs. transient ischaemic attack (TIA) (including timing), (iv) categorization of stroke into major and minor events based on the degree of disability as defined by conventional neurological assessment tools, and (v) sub-classification of strokes into haemorrhagic, ischaemic, and undetermined categories.

Table 3 outlines the diagnostic criteria and specific definitions for TIA and stroke as proposed by VARC. There is growing acceptance that neuroimaging is an important biomarker for the diagnosis of neuronal injury and stroke<sup>18,19</sup> and diffusion-weighted MRI is generally considered the procedure of choice in the context of acute neurological syndromes.<sup>30</sup> If a stroke is reported without evidence of confirmation of the diagnosis by the methods outlined in Table 3, the event may still be considered a stroke on the basis of the clinical presentation alone, but formal adjudication by qualified neurologists who are members of, or consultants to, a clinical events committee is mandatory. Patients with a global encephalopathy will not be reported as a stroke without unequivocal neuroimaging findings. Diagnosis of stroke in patients with a previously documented neurological deficit is more problematic and requires clinical assessment by a neurologist accompanied by appropriate new CT or MRI findings. In patients with a previous stroke and persistent neurological deficits, baseline pre-treatment neurological consultation and neuroimaging studies are recommended.

The earliest time of new neurological symptoms is defined as the *time of onset* of the stroke or TIA. When a patient awakens or begins responding (if previously unconscious) after the index

**Table 3 Stroke**

|   |
|---|
| <p>Stroke diagnostic criteria</p> <p>Rapid onset of a focal or global neurological deficit with at least one of the following: change in level of consciousness, hemiplegia, hemiparesis, numbness or sensory loss affecting one side of the body, dysphasia or aphasia, haemianopia, amaurosis fugax, or other neurological signs or symptoms consistent with stroke</p> <p>Duration of a focal or global neurological deficit <math>\geq 24</math> h; OR <math>&lt; 24</math> h, if therapeutic intervention(s) were performed (e.g. thrombolytic therapy or intracranial angioplasty); OR available neuroimaging documents a new haemorrhage or infarct; OR the neurological deficit results in death</p> <p>No other readily identifiable non-stroke cause for the clinical presentation (e.g. brain tumour, trauma, infection, hypoglycaemia, peripheral lesion, pharmacological influences)<sup>a</sup></p> <p>Confirmation of the diagnosis by at least one of the following:</p> <ul style="list-style-type: none"> <li>Neurology or neurosurgical specialist</li> <li>Neuroimaging procedure (MR or CT scan or cerebral angiography)</li> <li>Lumbar puncture (i.e. spinal fluid analysis diagnostic of intracranial haemorrhage)</li> </ul> <hr/> <p>Stroke definitions</p> <p>Transient ischaemic attack:</p> <ul style="list-style-type: none"> <li>New focal neurological deficit with rapid symptom resolution (usually 1–2 h), always within 24 h</li> <li>Neuroimaging without tissue injury</li> </ul> <p>Stroke: (diagnosis as above, preferably with positive neuroimaging study)</p> <ul style="list-style-type: none"> <li>Minor—Modified Rankin score <math>&lt; 2</math> at 30 and 90 days<sup>b</sup></li> <li>Major—Modified Rankin score <math>\geq 2</math> at 30 and 90 days</li> </ul> |
|---|

<sup>a</sup>Patients with non-focal global encephalopathy will not be reported as a stroke without unequivocal evidence based upon neuroimaging studies.

<sup>b</sup>Modified Rankin Score assessments should be made by qualified individuals according to a certification process. If there is discordance between the 30 and 90 day Modified Rankin Scores, a final determination of major vs. minor stroke will be adjudicated by the neurology members of the clinical events committee.

procedure with obvious new signs of a neurological deficit, the stroke or TIA is considered to have occurred during the index procedure. The diagnosis of a *transient ischemic attack* is defined as complete resolution of new neurological symptoms usually within 1–2 h but always within 24 h and also requires a normal neuroimaging study.<sup>18,19</sup> A stroke fulfilling the diagnostic criteria in *Table 3* is classified as a *major stroke* based upon ongoing significant clinical disability, defined as a Modified Rankin Score  $\geq 2$ . Although the initial Modified Rankin Score should be recorded after 7 days or at the time of hospital discharge, the attribution of clinically significant disability requires a Modified Rankin Score  $\geq 2$  at both 30 and 90 days follow-up (allowing sufficient time for stroke disability to stabilize). The Modified Rankin Score determinations should be performed by qualified individuals who have undergone a certification process.<sup>31–34</sup> A *minor stroke* must also fulfil stroke diagnostic criteria, with either resolution of new neurological symptoms within 24 h or persistence of symptoms  $> 24$  h and a Modified Rankin Score  $< 2$  at both 30 and 90 days follow-up. If there is discordance between the 30 and 90-day Modified Rankin Scores, the final determination of major vs.

**Table 4 Bleeding**

|  |
|--|
| <p>Life-threatening or disabling bleeding</p> <p>Fatal bleeding OR</p> <p>Bleeding in a critical area or organ, such as intracranial, intraspinal, intraocular, or pericardial necessitating pericardiocentesis, or intramuscular with compartment syndrome OR</p> <p>Bleeding causing hypovolemic shock or severe hypotension requiring vasopressors or surgery OR</p> <p>Overt source of bleeding with drop in haemoglobin of <math>\geq 5</math> g/dL or whole blood or packed red blood cells (RBCs) transfusion <math>\geq 4</math> units<sup>a</sup></p> <hr/> <p>Major bleeding</p> <p>Overt bleeding either associated with a drop in the haemoglobin level of at least 3.0 g/dL or requiring transfusion of two or three units of whole blood/RBC AND</p> <p>Does not meet criteria of life-threatening or disabling bleeding</p> <hr/> <p>Minor Bleeding</p> <p>Any bleeding worthy of clinical mention (e.g. access site haematoma) that does not qualify as life-threatening, disabling or major</p> |
|--|

<sup>a</sup>Given one *unit* of packed RBC typically will raise blood haemoglobin concentration by 1 g/dL, an estimated decrease in haemoglobin will be calculated.

minor strokes should be adjudicated by the neurology members of the clinical events committee. For the purposes of clinical trial endpoints, VARC advocates that only major strokes should be considered as an important safety endpoint, however, all neurological events should be reported as adverse events.

Stroke will be further stratified into ischaemic, haemorrhagic, or undetermined origin utilizing newly proposed definitions by an FDA consensus panel.<sup>35</sup> Ischaemic stroke is as an acute symptomatic episode of focal cerebral, spinal, or retinal dysfunction caused by an infarction of central nervous system tissue. Haemorrhagic stroke is an acute symptomatic episode of focal or global cerebral or spinal dysfunction caused by a non-traumatic intraparenchymal, intraventricular, or subarachnoid haemorrhage. An undetermined stroke is a stroke with insufficient information to allow categorization as either of ischaemic or haemorrhagic origin.

### Bleeding complications

Bleeding is a critical safety endpoint in evaluating contemporary pharmacological agents and interventional devices.<sup>36–44</sup> Valve Academic Research Consortium carefully reviewed several literature sources including (i) landmark clinical trials assessing the effects of anti-thrombotic medications in stable and acute coronary syndromes,<sup>45–59</sup> (ii) a report by the Control of Anticoagulation Subcommittee of the International Society on Thrombosis and Haemostasis,<sup>60</sup> and (iii) surgery guidelines after cardiac valve procedures.<sup>12</sup>

The definition of clinically meaningful bleeding was guided by the following principles: (i) the definition must be based on objective criteria, including an obvious source of bleeding or number of transfusions; (ii) serious or meaningful bleeding must result in death, be life-threatening, be proven to be associated with

increased long-term mortality, cause chronic sequelae, or consume major health-care resources. The VARC definition of bleeding complications (Table 4) is divided into life-threatening or disabling bleeding, major bleeding, and minor bleeding; anything but minor bleeding constitutes a serious or meaningful bleeding event.

Given the ample body of literature suggesting that administration of whole blood or red blood cell (RBC) transfusions in patients with cardiovascular pathology may be potentially harmful,<sup>41–44</sup> VARC considers that any whole blood or RBC transfusion needs to be reported in the case report forms, including the number of transfused units, regardless of the presence or absence of overt bleeding. Transfusions also need to be further stratified into those associated with overt bleeding and those in the absence of overt bleeding. Bleeding complications and transfusions should also be characterized relative to the time of occurrence including during the procedure, within the index hospitalization, or post-discharge.

## Acute kidney injury

The natural history of acute kidney injury (AKI) in a variety of clinical settings<sup>61–70</sup> is now well understood, including the recognition that even small decreases in kidney function can have a dramatic impact on the risk for subsequent mortality.<sup>69,70</sup> In recent reports, AKI has been observed in 12–28% of patients undergoing TAVI and was associated with a four times higher post-procedural mortality.<sup>71,72</sup>

In defining the stages of AKI, VARC proposes adopting serum creatinine criteria from the ‘modified’ RIFLE (Risk, Injury, Failure, Loss, and End-stage kidney disease) classification (Table 5).<sup>73</sup> The RIFLE classification<sup>74–76</sup> has been validated in the setting of intensive care units and cardiac surgery<sup>77–81</sup> and provides practical definitions for early stages of renal dysfunction when kidney injury can still be prevented, as well as stages when the kidney has already been damaged and renal failure is established. Modifications of the original RIFLE classification include two important changes: (i) smaller changes in serum creatinine (0.3 mg/dL) are included in stage 1 (‘Risk’)<sup>82</sup>; (ii) the ‘Loss’ and ‘End-stage kidney disease’

categories have been removed due to a lack of uniform indications and timing of renal replacement therapy (RRT) and variability in RRT resources in different countries. An outer bound of 72 h from the index procedure for diagnosing AKI was selected based on evidence that adverse outcomes were observed when the elevation occurred within 24 to 48 h of the procedure<sup>83</sup> and to ensure that the process was both acute and related to the procedure itself rather than as a consequence of post-procedure multi-organ system failure. Risk, Injury, Failure, Loss, and End-stage kidney disease classifications also stress the predictive value of urine output criteria in defining AKI, but VARC has not included this measure in the definition of AKI since urine outputs may not be measured accurately or routinely in all cases.

Valve Academic Research Consortium proposes to utilize the modified RIFLE classification to (i) capture even the earliest stages of AKI (stage 1) on case report forms, (ii) define AKI as either stage 2 or 3, and (iii) report any case of RRT (haemodialysis, peritoneal dialysis, or haemofiltration) occurring during the index hospitalization or within 30 days after the index procedure. Given the well-recognized damaging impact of contrast media on renal function, VARC also recommends to report the volume and type of contrast medium used during the index procedure.

## Vascular complications

Recent TAVI literature indicates that major vascular complications using various non-standardized definitions (e.g. with or without including the need for blood transfusions) occur at a frequency of 4–34% and are associated with a two- or three-fold higher 30-day mortality.<sup>84–87</sup> In defining vascular complications, VARC referenced the reporting standards of the Society of Vascular Surgery for defining and reporting vascular complications following endovascular aortic graft repair procedures.<sup>88</sup>

Valve Academic Research Consortium proposes to report both major and minor vascular complications, but to only consider major vascular complications as an important clinical endpoint. Of note, the ‘access site’ is defined as any location (arterial or venous) traversed by a guide-wire, a catheter or a sheath [including the left ventricular (LV) apex and the aorta] and ‘access related’ is defined as any adverse clinical consequence possibly associated with any of the access sites used during the procedure. The VARC definitions for major and minor vascular complications are described in Table 6.

Many vascular situations require special notice. Femoral vascular access and closure in many centers is routinely achieved using surgical cut-down procedures, and therefore, pre-planned surgical access and/or closure should be considered as part of the procedure and not as a complication. Similarly, uncomplicated non-femoral (e.g. retroperitoneal, iliac, subclavian, or aortic) surgical access for sheath entry (planned or unplanned) is not considered a vascular complication, unless untoward clinical consequences are documented (e.g. bleeding complications). However, interventional or surgical repair for failed percutaneous closure of the arteriotomy site during the index procedure without other clinical sequelae (Table 6) is considered a minor vascular complication. Considering the recent proliferation of vascular access approaches and the recognition that specific access sites and techniques may be associated with either increased or decreased complications,

**Table 5 Acute kidney injury (modified RIFLE classification)**

### Change in serum creatinine (up to 72 h) compared with baseline

|                      |   |
|----------------------|---|
| Stage 1              | Increase in serum creatinine to 150–200% (1.5–2.0 × increase compared with baseline) or increase of ≥0.3 mg/dL (≥26.4 μmol/L)   |
| Stage 2              | Increase in serum creatinine to 200–300% (2.0–3.0 × increase compared with baseline) or increase between >0.3 mg/dL (>26.4 μmol/L) and <4.0 mg/dL (<354 μmol/L)                       |
| Stage 3 <sup>a</sup> | Increase in serum creatinine to ≥300% (>3 × increase compared with baseline) or serum creatinine of ≥4.0 mg/dL (≥354 μmol/L) with an acute increase of at least 0.5 mg/dL (44 μmol/L) |

<sup>a</sup>Patients receiving renal replacement therapy are considered to meet Stage 3 criteria irrespective of other criteria.

**Table 6** Vascular access site and access-related complications

|  |
|--|
| Major vascular complications   |
| Any thoracic aortic dissection   |
| Access site or access-related vascular injury (dissection, stenosis, perforation, rupture, arterio-venous fistula, pseudoaneurysm, haematoma, irreversible nerve injury, or compartment syndrome) leading to either death, need for significant blood transfusions ( $\geq 4$ units), unplanned percutaneous or surgical intervention, or irreversible end-organ damage (e.g. hypogastric artery occlusion causing visceral ischaemia or spinal artery injury causing neurological impairment) |
| Distal embolization (non-cerebral) from a vascular source requiring surgery or resulting in amputation or irreversible end-organ damage  |
| Minor vascular complications   |
| Access site or access-related vascular injury (dissection, stenosis, perforation, rupture, arterio-venous fistula or pseudoaneurysms requiring compression or thrombin injection therapy, or haematomas requiring transfusion of $\geq 2$ but $< 4$ units) not requiring unplanned percutaneous or surgical intervention and not resulting in irreversible end-organ damage  |
| Distal embolization treated with embolectomy and/or thrombectomy and not resulting in amputation or irreversible end-organ damage  |
| Failure of percutaneous access site closure resulting in interventional (e.g. stent-graft) or surgical correction and not associated with death, need for significant blood transfusions ( $\geq 4$ units), or irreversible end-organ damage   |

**Table 7** Potential failure modes of prosthetic valve dysfunction

|   |
|---|
| Aortic stenosis   |
| Stent creep   |
| Pannus  |
| Calcification   |
| Support structure deformation (out-of-round configuration), under-expansion, fracture, or trauma (cardio-pulmonary resuscitation, blunt chest trauma)                                       |
| Mal-sizing (prosthesis-patient mismatch)  |
| Endocarditis  |
| Prosthetic valve thrombosis   |
| Native leaflet prolapse impeding prosthetic leaflet motion  |
| Aortic regurgitation  |
| Pannus  |
| Calcification   |
| Support structure deformation (out-of-round configuration), recoil, under-expansion, fracture, insufficient radial strength, or trauma (cardio-pulmonary resuscitation, blunt chest trauma) |
| Endocarditis  |
| Prosthetic valve thrombosis   |
| Malposition (too high, too low)   |
| Acute mal-coaptation  |
| Leaflet wear, tear/perforation, prolapse, or retraction   |
| Suture breakage or disruption   |
| Native leaflet prolapse impeding prosthetic leaflet motion  |

VARC strongly recommends that detailed information is recorded on the access site and technique for each procedure.

A special circumstance relates to complications associated with the left-ventricular apex site during transapical TAVI procedures. Although such complications are less frequent than transarterial vascular complications, a recent report indicates that the clinical consequences (including death) can be more serious.<sup>89</sup> Major complications associated with transapical TAVI procedures include bleeding, pseudoaneurysm formation (with or without rupture), and haemodynamic instability requiring urgent transarterial cardio-pulmonary bypass support. Valve Academic Research Consortium proposes that all such complications associated with transapical TAVI be reported in case report forms and that clinical consequences resulting from such complications be registered under the appropriate clinical endpoints (e.g. bleeding, stroke, mortality).

### Prosthetic valve performance

The clinical presentation of patients with prosthetic valve dysfunction is usually consistent with symptoms and signs of either valvular regurgitation or stenosis. Valve Academic Research Consortium proposes only two criteria to evaluate impaired prosthetic valve performance: (i) prosthetic valve haemodynamics assessed by echocardiography and (ii) associated clinical findings indicating impaired cardiovascular or valvular function (e.g. new or worsening congestive heart failure). Transthoracic echocardiography (TTE) is

usually adequate to evaluate prosthetic aortic valve function,<sup>90</sup> although transoesophageal echocardiography (TEE) may be very useful in the setting of technically challenging or complex cases. Serial echocardiography evaluations after surgical AVR and TAVI should be performed at baseline, soon after the index procedure (ideally within 24–48 h, but always before discharge), at 1 month (especially for TAVI), 12 months, and yearly thereafter.<sup>91,92</sup> This follow-up schedule is more intensive than recommended in the AHA/ACC and ESC guidelines for follow-up after surgical AVR,<sup>91,92</sup> but more frequent documentation of valve function and position is considered desirable for TAVI clinical trials.

Although the VARC definitions for impaired prosthetic valve performance discount mechanistic characterizations, valve failure mode(s) should be recorded whenever possible in case report forms (Table 7). In addition to echocardiography, multi-slice computed tomography may also provide useful insights into the responsible patho-biological mechanisms of device malfunction.<sup>93,94</sup>

### Prosthetic aortic stenosis and regurgitation

Utilizing the recent prosthetic valve echocardiography guidelines,<sup>90</sup> the severity of prosthetic aortic valve stenosis is graded as (i) normal, (ii) possible, or (iii) significant (Table 8) and prosthetic aortic valve regurgitation (central or paravalvular) as (i) mild, (ii) moderate, or (iii) severe (Table 9). The clinical significance of prosthetic valve dysfunction is further supported by the presence of

**Table 8** Prosthetic aortic valve stenosis criteria<sup>a</sup>

| Parameter  | Normal                    | Possible stenosis          | Significant stenosis         |
|--|---------------------------|----------------------------|------------------------------|
| Peak velocity (m/s) <sup>b</sup>                         | <3                        | 3–4                        | >4                           |
| Mean gradient (mmHg) <sup>b</sup>                        | <20                       | 20–35                      | >35                          |
| Doppler velocity index                                   | ≥0.30                     | 0.29–0.25                  | <0.25                        |
| Effective orifice area (cm <sup>2</sup> )                | >1.2                      | 1.2–0.8                    | <0.80                        |
| Contour of the jet velocity through the prosthetic valve | Triangular, early peaking | Triangular to intermediate | Rounded, symmetrical contour |
| Acceleration time (ms)                                   | <80                       | 80–100                     | >100                         |

<sup>a</sup>In conditions of normal or near normal stroke volume (50–70 mL).

<sup>b</sup>These parameters are more affected by flow, including concomitant aortic regurgitation.

**Table 9** Prosthetic aortic valve regurgitation criteria (central and paravalvular)

| Parameter  | Mild                            | Moderate              | Severe                   |
|--|---------------------------------|-----------------------|--------------------------|
| Valve structure and motion                                     |                                 |                       |                          |
| Mechanical or bioprosthetic                                    | Usually normal                  | Usually abnormal      | Usually abnormal         |
| Structural parameters  |                                 |                       |                          |
| Left ventricular size  | Normal                          | Normal/mildly dilated | Dilated                  |
| Doppler parameters (qualitative or semi-quantitative)          |                                 |                       |                          |
| Jet width in central jets (% LVO diameter): color <sup>a</sup> | Narrow (≤25%)                   | Intermediate (26–64%) | Large (≥65%)             |
| Jet density: CW Doppler  | Incomplete or faint             | Dense                 | Dense                    |
| Jet deceleration rate (PHT, ms): CW Doppler <sup>b</sup>       | Slow (>500)                     | Variable (200–500)    | Steep (<200)             |
| LV outflow vs. pulmonary flow: PW Doppler                      | Slightly increased              | Intermediate          | Greatly increased        |
| Diastolic flow reversal in the descending aorta:               |                                 |                       |                          |
| PW Doppler   | Absent or brief early diastolic | Intermediate          | Prominent, holodiastolic |
| Circumferential extent of paraprosthetic AR (%) <sup>c</sup>   | <10                             | 10–20                 | >20                      |
| Doppler parameters (quantitative)                              |                                 |                       |                          |
| Regurgitant volume (mL/beat)                                   | <30                             | 30–59                 | >60                      |
| Regurgitant fraction (%)                                       | <30                             | 30–50                 | >50                      |

<sup>a</sup>Parameter applicable to central jets and is less accurate in eccentric jets.

<sup>b</sup>Influenced by left ventricular compliance.

<sup>c</sup>For paravalvular aortic regurgitation.

AR, aortic regurgitation; CW, continuous wave; LVO, left ventricular outflow; PW, pulsed wave.

clinical signs, symptoms, and/or events (e.g. re-hospitalization for worsening symptoms, re-operation or death).

Transcatheter aortic valve implantation devices are associated with a higher frequency of mild and moderate paravalvular aortic regurgitation (AR) than surgical AVR.<sup>95–107</sup> There is a need to develop improved definitions and to better understand the long-term clinical implications of paravalvular prosthetic AR. Unfortunately, the precise grading of paravalvular AR remains controversial and many echocardiography experts believe that grading schemes for prosthetic central and paravalvular AR should be different.<sup>90,108–111</sup> Recently, in the setting of TAVI, criteria for assessing paravalvular AR severity have emphasized a ‘jet anatomy’ classification, stressing the location, circumferential extent, and width of the AR jet.<sup>90</sup> Haemodynamic factors can also be useful to assess the AR severity immediately after valve implantation (e.g. loss of aortic diastolic notch, equalization of end-diastolic aortic and left ventricular pressures). Since there is lack of clarity concerning

the long-term clinical implications of mild and moderate paravalvular AR after TAVI, echocardiography core laboratories are useful to ensure consistent evaluation methods. Echocardiograms should be performed annually in those patients known to have post-procedural paravalvular AR.

### Prosthetic aortic valve thrombosis and endocarditis

Although prosthetic valve thrombosis and prosthetic valve endocarditis have been included as potential failure modes for prosthetic valve dysfunction (Table 7), they require reporting as individual endpoints. Valve thrombosis is any thrombus attached to or near an implanted valve that occludes part of the blood flow path, interferes with valve function, or is sufficiently large to warrant treatment.<sup>12</sup> Furthermore, valve thrombus found at autopsy in a patient whose cause of death was not valve related or found at operation for an unrelated indication should also be reported as

valve thrombosis. The diagnosis of prosthetic valve thrombosis is best discerned during an echocardiographical examination or during surgical exploration. There have already been case reports and anecdotes<sup>112</sup> of transcatheter prosthetic valve thrombosis with and without important clinical consequences.


The diagnosis of prosthetic valve endocarditis is based on one of the following criteria<sup>12</sup>:

- reoperation with evidence of abscess, paravalvular leak, pus, or vegetation confirmed as secondary to infection by histological or bacteriological studies;
- autopsy findings of abscess, pus, or vegetation involving a repaired or replaced valve;
- in the absence of reoperation or autopsy, fulfilling the Duke Criteria for endocarditis.<sup>113</sup>

Isolated case reports of transcatheter aortic valve endocarditis have already been published.<sup>114,115</sup> Owing to the variability in transcatheter valve designs and positioning within the aortic root, meticulous reporting of the pattern of endocarditis is mandatory.<sup>116</sup>

## Prosthetic valve ‘associated’ complications

Depending on the design characteristics and final implant position, prosthetic aortic valves may come in close contact with the anterior mitral valve leaflet, the intervalvular fibrosa, the aortic annulus, the ventricular septum, the aortic sinuses and root, the coronary arteries, and the cardiac conduction system. Collectively, these anatomic structures, which are contiguous with the prosthetic aortic valve, are referred to as the *aortic valvar complex* (Figure 2). As such, prosthetic aortic valve procedures, and in particular TAVI, may have untoward effects on any of these structures which may result in important clinical consequences. Therefore, VARC



**Figure 2** The aortic valvar complex, including the aortic valve, annulus, sinuses, aorta, coronary arteries, membranous septum, and the mitral valve.

proposes to group these complications as a separate endpoint category. However, it must be noted, that some of these adverse events may not be directly related to The Valve prosthesis itself, but may occur before or after valve implantation (e.g. conduction disturbances after pre-implant balloon aortic valvuloplasty).

## Conduction disturbances and cardiac arrhythmias

The close anatomical relationship between the aortic valvar complex and the branching atrioventricular bundle explains the possible development of conduction abnormalities following prosthetic aortic valve procedures.<sup>117,118</sup> Following surgical AVR, new-onset bundle branch block has been reported in 16–32% of patients and the need for permanent pacemakers in 3–8% of patients.<sup>119–123</sup> In early experiences with TAVI, new-onset bundle branch block has occurred in up to 45% of patients and the need for permanent pacemakers has varied from as low as 4% to as high as 33%.<sup>95–98,124–128</sup> Differences among devices and heterogeneity in physician and country-based healthcare thresholds may explain the significant inter-hospital variability in new permanent pacemaker requirements after TAVI.

Although the implications of persistent left bundle branch block (LBBB) after TAVI are currently unknown, the presence of new bundle branch block after surgical AVR has been associated with increased risk of subsequent arrhythmic events during follow-up (specifically, syncope, AV dissociation, and sudden death).<sup>119,120</sup> Owing to this association between conduction system abnormalities and adverse patient outcomes following surgical AVR and several anecdotal reports after TAVI of either early post-discharge severe bradyarrhythmic events or sudden cardiac death,<sup>15</sup> VARC recommends to carefully document the occurrence of new conduction system abnormalities (left bundle branch block and third degree atrioventricular block), as well as the requirements and indications for new permanent pacemakers within 30 days after the procedure. The timing (days) and location (intra-procedural, in-hospital, or post-discharge) of the event should also be recorded. To accurately capture such events, daily ECGs and continuous telemetry ECG monitoring should be considered while the patients are in-hospital, and should be required in patients with any evidence of new conduction abnormalities or arrhythmias.

Although conduction abnormalities associated with TAVI have been a recent concern, it bears noting that new onset atrial fibrillation and ventricular arrhythmias have also been observed after both TAVI and surgical AVR procedures.<sup>128</sup> In particular, new onset atrial fibrillation occurs in as many as 20–30% of patients after conventional surgical AVR<sup>129,130</sup> and any valid comparison of transcatheter vs. surgical aortic valve treatment strategies should include a careful analysis of post-therapy supra-ventricular and ventricular arrhythmias.

## Coronary obstruction

Mechanical coronary artery obstruction following TAVI or surgical AVR is rare and occurs in <1% of patients.<sup>96,97</sup> The obstruction typically occurs during the index procedure. Importantly, clinical signs and symptoms may be subtle and not appreciated until

after the procedure. Possible mechanisms for mechanical coronary obstruction include (i) impingement of the coronary ostia by the Valve support structure in the setting of suboptimal valve positioning and/or 'small aortic root' anatomy; (ii) embolization from calcium, thrombus, air, or endocarditis; (iii) displacement of native aortic valve leaflets towards the coronary ostia during TAVI,<sup>131,132</sup> and (iv) suture-related kinking or obstruction or cannulation-related obstruction of the coronary ostia associated with surgical AVR.

The diagnosis of TAVI-associated coronary obstruction can be determined by imaging studies (coronary angiography, intravascular ultrasound, multi-slice CT angiography, or echocardiography), surgical exploration, or autopsy findings. Cardiac biomarker elevations and ECG changes indicating new ischaemia provide corroborative evidence.

### Other prosthesis-related adverse events

The short- and long-term consequences of contact, trauma, or impingement on the anterior mitral valve leaflet by the ventricular end of a transcatheter aortic valve are currently unknown. Nevertheless, any new mitral valve dysfunction (e.g. worsening mitral regurgitation or stenosis) or disruption (e.g. chordal rupture, leaflet perforation, anterior mitral valve leaflet aneurysm) related to contact with the transcatheter valve implant or mitral valve endocarditis<sup>114–116</sup> should be carefully documented. Other infrequent complications following TAVI include new ventricular septal defects and aortic root rupture/perforation/dissection, occurring either during the pre-implant balloon aortic valvuloplasty, or during the transcatheter valve implant.<sup>133,134</sup>

### Clinical benefit endpoints

In addition to the avoidance of mortality, specific endpoints to establish the clinical benefit after TAVI are important. Objective benefit parameters derived from the heart failure literature can be adapted in valve-related clinical trials.<sup>135</sup> Several choices are available, including exercise performance,<sup>136</sup> assessment of New York Heart Association (NYHA) functional status,<sup>137</sup> and various quality of life<sup>138</sup> and frailty questionnaires.<sup>139</sup> Each of these symptom evaluation tools has strengths and weaknesses in the TAVI patient population, which is disproportionately represented by elderly, frail, individuals with multiple co-morbidities. For instance, exercise test performance is an appealing endpoint, but as aortic valve therapy studies are unblinded, they may be biased and they can be difficult to perform in high-risk TAVI patients.

Valve Academic Research Consortium has also considered a categorical endpoint of clinical benefit which captures failure of current AS therapy; *hospitalization for symptoms of cardiac or valve-related decompensation*, at least 30 days after the index procedure (surgical AVR or TAVI). This endpoint mandates careful adjudication by a clinical events committee and is defined as hospitalization for symptoms of valve or cardiac deterioration (e.g. new or worsening heart failure, angina, or syncope) requiring either a valve procedure (surgery or interventional treatment) or intensification of medical management (new or increased use of inotropes, vasopressors, diuretics, and/or vasodilators).

Quality-of-life and healthcare economic instruments can be useful to assess disability and impairment due to congestive heart failure (e.g. Kansas City Cardiomyopathy Questionnaire)<sup>140</sup> and for mapping health status compared with population-level utility weights (e.g. EuroQOL questionnaire).<sup>141–143</sup> However, quality-of-life questionnaires are also prone to bias and must be uniformly administered. The time points for assessment of the aforementioned clinical benefit endpoints should be at 30 days, at 6 months, and at 1 year after initiating therapy. Valve Academic Research Consortium recommends that if any measure of clinical benefit is utilized in clinical trials, there must be careful oversight and adjudication by experienced clinical events committees.

The assessment of 'frailty' in patients with advanced valvular heart disease has become increasingly important and is usually not included in surgical-risk algorithms. Frailty is loosely defined as a biological syndrome of decreased reserve and resistance to stressors, resulting from cumulative declines across multiple physiological systems, and causing vulnerability to adverse outcomes.<sup>139</sup> Various frailty indices have been developed and have been correlated with worsening clinical outcomes in geriatric patients in intensive care units and after surgery.<sup>144–146</sup> In general, the evaluation of frailty demands a composite analysis of several categorical and continuous variables including mobility, strength, endurance, activities of daily living, cognitive impairment, and nutritional status (as discerned by body mass index and biomarkers such as serum albumin).<sup>147,148</sup> Although there is no standard frailty index which has been applied and validated in high-risk AS patients, multiple preliminary efforts are ongoing and VARC proposes to include measures of frailty in future clinical trials as a component of clinical benefit endpoints.

### Therapy-specific endpoints

Given the complex nature of TAVI procedures and the rapid evolution of devices and procedural techniques, VARC proposes to record in case report forms (but not as formal endpoints) an open category of therapy-specific endpoints which may be relevant to clinical outcomes or device performance. Examples of such events include the following: (i) the unplanned use of cardiopulmonary bypass to manage haemodynamic compromise or to reverse procedural complications; (ii) conversion from a 'failed' percutaneous transcatheter procedure to an 'open' surgical AVR or to a surgical-access TAVI<sup>149,150</sup>; (iii) ventricular perforation (for any reason) with and without cardiac tamponade<sup>151</sup>; (iv) prosthetic valve migration or dislocation from the native aortic valve landing zone<sup>152,153</sup>; (v) frequency, reasons, and results of post-TAVI balloon dilation; (vi) frequency, reasons, and results after placement of a second valve over the original valve, so-called TAVI 'valve-in-valve'<sup>154,155</sup>; (vii) integrity of the support structure, including strut fractures, compression or other evidence of geometry distortion (requires careful serial imaging modalities including cine-fluoroscopic analyses and echocardiography)<sup>90,156–158</sup>; (viii) instances of device recapture (with or without repositioning), or retrieval (removal from the body) which occur during the index procedure; (ix) re-intervention (either percutaneous or surgical) for any reason after the index procedure.<sup>159</sup> As appropriate, the timing of these events (during the index procedure, in-hospital,

**Table 10** Composite endpoints

|  |
|--|
| Device success   |
| Successful vascular access, delivery and deployment of the device and successful retrieval of the delivery system  |
| Correct position of the device in the proper anatomical location   |
| Intended performance of the prosthetic heart valve (aortic valve area $> 1.2 \text{ cm}^2$ and mean aortic valve gradient $< 20 \text{ mmHg}$ or peak velocity $< 3 \text{ m/s}$ , without moderate or severe prosthetic valve AR) |
| Only one valve implanted in the proper anatomical location   |
| Combined safety endpoint (at 30 days)  |
| All-cause mortality  |
| Major stroke   |
| Life-threatening (or disabling) bleeding   |
| Acute kidney injury—Stage 3 (including renal replacement therapy)  |
| Peri-procedural MI   |
| Major vascular complication  |
| Repeat procedure for valve-related dysfunction (surgical or interventional therapy)  |
| Combined efficacy endpoint (at 1 year or longer)   |
| All-cause mortality (after 30 days)  |
| Failure of current therapy for AS, requiring hospitalization for symptoms of valve-related or cardiac decompensation   |
| Prosthetic heart valve dysfunction (aortic valve area $< 1.2 \text{ cm}^2$ and mean aortic valve gradient $\geq 20 \text{ mmHg}$ or peak velocity $\geq 3 \text{ m/s}$ , OR moderate or severe prosthetic valve AR)                |

or post-discharge) should be carefully recorded. This category is intended to be a dynamic platform and should be added to the case report forms.

## Clinically relevant composite endpoints

Although VARC discourages the overuse of composite endpoints, to achieve overall impressions of safety and effectiveness may require the incorporation of more than single endpoints. These strategic assessments of TAVI as an alternative therapy should ideally include device, procedure, and patient-oriented factors. Valve Academic Research Consortium proposes three composite endpoints (Table 10): device success (intra-procedure), a combined safety endpoint (at 30 days), and a combined efficacy endpoint (at 1 year or longer).

*Device success* is a ‘technical’ composite endpoint meant to characterize the acute device and procedural factors which underlie vascular access, delivery, and performance of the TAVI system. Echocardiography should be routinely utilized as the standard for measuring prosthetic valve stenosis and regurgitation immediately after TAVI, and should always be performed in a resting state, either within 24–48 h after the index procedure or before hospital discharge.

The 30-day *combined safety endpoint* is a hierarchical composite of the most relevant patient-oriented safety endpoints previously defined by VARC (Table 10). In addition, a repeat procedure in

the first 30 days (either surgery or intervention) to treat valve-related dysfunction is also incorporated in this endpoint. Examples of urgent repeat procedures would include balloon aortic valvuloplasty or repeat TAVI (valve-in-valve) to treat either paravalvular or central severe AR after the TAVI. The focus on 30-day events after the index procedure is meant to isolate safety concerns largely pertaining to early device performance and the procedure. Nonetheless, overall patient safety also requires a careful examination of pertinent individual safety endpoints over the life history of the device.

The time-sensitive assessment of TAVI effectiveness requires a more delayed *combined efficacy endpoint* incorporating major clinical and valve performance factors. Valve Academic Research Consortium proposes a 1-year (or longer) time interval for the combined efficacy endpoint integrating three important endpoints: (i) all-cause mortality *after 30 days*, meant to reflect therapy effectiveness by measuring prevention of AS-related mortality over time; (ii) failure of the current therapy for AS, requiring hospitalization for symptoms of valve-related or cardiac decompensation (adjudicated episodes of heart failure, angina, or syncope requiring an aortic valve procedure or intensification of medical management); (iii) evidence of prosthetic valve dysfunction, defined using strict echocardiography criteria, possibly in conjunction with other signs of functional deterioration.

## Discussion

The VARC was convened in response to an urgent call for standardized clinical research processes involving the emerging field of transcatheter valve therapies, and more specifically, TAVI in high-surgical-risk patients with AS. The inter-disciplinary nature of TAVI, combining aspects of both surgical and interventional therapies, presented special challenges and required an enlightened and collaborative approach to the development of clinical research recommendations and endpoint definitions.<sup>160,161</sup> The VARC initiative is an attempt to achieve a necessary consensus among the various subspecialties and stakeholders, such that this innovative treatment strategy may be evaluated objectively and according to a set of practical endpoint definitions.

This consensus manuscript is not intended to be interpreted as a ‘guidelines’ or ‘guidance’ document and although thoroughly reviewed by individuals from seven cardiology and cardiac surgery societies, the content has not been subjected to a formal society guidelines review process. These standardized endpoints are measurable, apply to both predicate surgical and new transcatheter therapies, can be adjudicated by clinical events committees, and can be used to compare findings from different clinical trials. By intent, this consensus manuscript was not device-specific and the definitions can be applied to next generation and iterative TAVI devices already under early stages of clinical investigation.<sup>162–165</sup>

Given the rapid growth in transcatheter valve therapies, and the potential exposure of this technology to lower risk patient populations, it is certain that this preliminary attempt to arrive at consensus endpoint definitions for TAVI will need refinement in the future. In principle, the consensus process calls for the highest standards of clinical research, including (i) inter-disciplinary experts gathering to arrive at standardized endpoint definitions,

(ii) harmonized and well-structured data collection, interpretation, and reporting for specific TAVI-related clinical events, and (iii) the consistent use of central core laboratories and independent, blinded endpoint adjudication.

Many of the endpoints discussed in this manuscript are sufficiently general that they can be applied to other AS populations and to other valvular heart disease clinical research scenarios, both surgical and interventional. This is particularly germane to TAVI clinical research, as new studies involving lower risk AS patients are already being considered. Importantly, recent reports and randomized trials using new catheter-based *mitral valve* therapies to treat mitral regurgitation<sup>166–168</sup> also suffer from non-standardized endpoint definitions and might well benefit from a comparable VARC consensus effort.

This consensus manuscript, which represents the ‘first step’ in a much longer road to help improve clinical research in valvular heart disease, has several limitations. The endpoint definitions were intended to be reasonably broad, but nonetheless in some instances are also intentionally narrow to address the specific considerations of TAVI in high-surgical-risk patients with severe AS. Therefore, application of all of these endpoint definitions to other patient populations may be problematic. The important area of pre-clinical device testing, both assessments of valve and support structure properties and *in vivo* animal studies, is beyond the scope of this manuscript. Other aspects of clinical trial design and clinical trial methodologies are also essential to optimize clinical research, but similarly, a comprehensive treatment of these subjects could not be included in this manuscript. Finally, many global endpoints, such as stroke and bleeding and some specific endpoints, such as paravalvular regurgitation, are themselves in a state of evolution, subject to modifications by other consensus committees in the near future.

The VARC process embodied in this manuscript was an ambitious multi-disciplinary attempt to bring order through consensus, thereby providing standardization of clinical research in the burgeoning area of transcatheter aortic valve therapy. Hopefully, this template can also serve as a model to improve clinical research methodologies in the evaluation of new therapies for other cardiovascular diseases.

## Funding

Grants were provided to the ARC Board including representatives of The Cardiovascular Research Foundation, Cardialysis, Duke Clinical Research Institute and Harvard Clinical Research Institute to cover the costs of travel, meeting rooms, and lodging for academic attendees at the San Francisco and Amsterdam meetings by Edwards Lifesciences and Medtronic Corporation. All funds not utilized for the aforementioned travel-related purposes have been returned to the sponsors. Funding to pay the Open Access publication charges for this article was provided by Cardialysis BV on behalf of the Valve Academic Research Consortium.

**Conflict of interest:** VARC Participants will provide Conflict of Interest Disclosures individually prior to publication. The VARC meetings involved members of the Interventional Cardiology Devices Branch, of the Office of Device Evaluation, Center for Devices and Radiological Health, USFDA. The opinions or assertions herein are the private views of the authors and are not to be construed as reflecting the views of the FDA.

## Appendix 1

### Valve Academic Research Consortium Contributing Groups

#### (1) Academic Research Organizations

Cardialysis (Rotterdam, the Netherlands)  
 Cardiovascular Research Foundation (New York, NY, USA)  
 Duke Clinical Research Institute (Durham, NC, USA)  
 Harvard Clinical Research Institute (Boston, MA, USA)

#### (2) Societies

American Association for Thoracic Surgery  
 American College of Cardiology  
 American Heart Association  
 European Association for CardioThoracic Surgery  
 European Society of Cardiology  
 Society of Cardiac Angiography and Intervention  
 Society of Thoracic Surgeons.

#### (3) US Food and Drug Administration

#### (4) Industry Representatives

## Appendix 2

### Valve Academic Research Consortium participants

#### (1) Clinical Research Organizations

##### (1) Cardialysis/Erasmus MC—Rotterdam, the Netherlands

Morel M.A.  
 Piazza, N.  
 Serruys, P.W.  
 Van Es, G.A.  
 Van Mieghem, N.  
 Vranckx, P.

##### (2) Cardiovascular Research Foundation—New York City, NY, USA

Caixeta, A.  
 Dalton, K.  
 Haratani, N.  
 Kirtani, A.  
 Kodali, S.  
 Lansky, A.  
 Leon, M.B.  
 Mehran, R.  
 Nikolsky, E.  
 Williams, M.

##### (3) Duke Clinical Research Institute—Durham, NC, USA

Krucoff, M.W.  
 Petersen, J.

##### (4) Harvard Cardiovascular Research Institute—Boston, MA, USA

Cutlip, D.E.  
 Pinto, D.

## (2) Physician Society Representatives and Experts

## Surgeons

Adams, D.—Mt. Sinai Medical Center, New York City, USA

Kappetein, A.P.—Erasmus MC, Rotterdam, the Netherlands

Mack, M.—Medical City Hospital, Dallas, TX, USA

Miller, C.—Stanford University, CA, USA

Mohr, F.—Herzzentrum Universität Leipzig, Leipzig, Germany

Nataf, P.—CHU Bichat, Paris, France

Smith, C.—Columbia University Medical Center, New York City, NY, USA

Verrier, E.—University of Washington, Seattle, WA, USA

## Cardiologists

Bailey, S.—University of Texas, San Antonio, TX, USA

Bonan, R.—Montreal Heart Institute, Montreal, Canada

Bonhoeffer, P.—Great Ormond Street Hospital for Children, London, UK

Bonow, R.—Northwestern University, Chicago, IL, USA

Douglas, P.—Duke University, Durham, NC, USA

Gillam, L.—Columbia University Medical Center, New York City, NY, USA

Iung, B.—CHU Bichat, Paris, France

Popma, J.J.—Beth Israel—Deaconess Medical Center, Boston, MA, USA

Thomas, M.—Guys and St Thomas Hospital, London, UK

Tuzcu, M.—Cleveland Clinic Foundation, Cleveland, OH, USA

Vahanian, A.—CHU Bichat, Paris, France

Virmani, R.—CV Pathology Institute, Gaithersburg, MD, USA

Webb, J.G.—St Paul's Hospital, Vancouver, BC, Canada

Windecker, S.—University Hospital Bern, Bern, Switzerland

## Biostatistics/Clinical Epidemiology

Blackstone, E.—Cleveland Clinic Foundation, Cleveland, OH, USA

Pocock, S.—London School of Hygiene and Tropical Medicine, London, UK

Takkenberg, J.J.M.—Erasmus MC, Rotterdam, the Netherlands

US Food and Drug Administration

Hillebrenner, M.

Swain, J.

Zuckerman, B.

Medicines and Healthcare products Regulatory Agency (MHRA)

Ludgate, S.

## Industry Representatives

Akin, J.—Edwards Lifesciences, Orange, CA, USA

Armitage, T.—Medtronic, Minneapolis, MN, USA

Martin, K.—Sadra Medical, Los Gatos, CA, USA

Sheahan, B.—Direct Flow Medical, Santa Rosa, CA, USA

Simonton, C.—Abbott Vascular, Santa Clara, CA, USA

Wilson, R.—Heart Leaflet Technologies, Maple Grove, MN, USA

## Appendix 3

### Minimum data collection and endpoint requirements after TAVI

- Mortality (all cause and cardiovascular)
- Myocardial Infarction (peri-procedural and spontaneous)
- Stroke (major and minor)
- Bleeding (life threatening or disabling and major)
- Acute kidney injury (modified RIFLE stage 2 and 3, including RRT)
- Vascular complications (major)
- Prosthetic valve performance (requires serial echocardiography assessments)
  - (a) Prosthetic valve stenosis (possible and significant) and regurgitation (moderate or severe with special reference to paravalvular regurgitation)
  - (b) Prosthetic valve thrombosis
  - (c) Prosthetic valve endocarditis
- Prosthetic valve-associated complications
  - (a) Conduction disturbances and cardiac arrhythmias (including new LBBB, new permanent pacemaker implantation, and new supra-ventricular or ventricular arrhythmias) and
  - (b) Coronary obstruction
- Clinical benefit endpoints
  - (a) Symptom status (global assessments using NYHA classification and some measure of quality of life)
  - (b) Repeat hospitalization (>30 days after the index procedure) for valve-related or cardiac decompensation)
- Therapy specific endpoints (ventricular perforation at any time resulting in cardiac tamponade, prosthetic valve embolization, and acute or delayed valve-in-valve treatment)
- Composite endpoints
  - (a) Device success
  - (b) Combined safety endpoint (at 30 days)
  - (c) Combined efficacy endpoint (at 1 year or longer)

### References

1. Cribier A, Eltchaninoff H, Bash A, Borenstein N, Tron C, Bauer F, Derumeaux G, Anselme F, Laborde F, Leon MB. Percutaneous transcatheter implantation of an aortic valve prosthesis for calcific aortic stenosis: first human case description. *Circulation* 2002;**106**:3006–3008.
2. Cribier A, Eltchaninoff H, Tron C, Bauer F, Agatiello C, Sebagh L, Bash A, Nusimovici D, Litzler PY, Bessou JP, Leon MB. Early experience with percutaneous transcatheter implantation of heart valve prosthesis for the treatment of end-stage inoperable patients with calcific aortic stenosis. *J Am Coll Cardiol* 2004;**43**:698–703.
3. Iung B, Baron G, Butchart EG, Delahaye F, Gohlke-Barwolf C, Levang OW, Tornos P, Vanoverschelde JL, Vermeer F, Boersma E, Ravaud P, Vahanian A. A prospective survey of patients with valvular heart disease in Europe: The Euro Heart Survey on Valvular Heart Disease. *Eur Heart J* 2003;**24**:1231–1243.
4. Stuge O, Liddicoat J. Emerging opportunities for cardiac surgeons within structural heart disease. *J Thorac Cardiovasc Surg* 2006;**132**:1258–1261.
5. Varadarajan P, Kapoor N, Bansal RC, Pai RG. Clinical profile and natural history of 453 nonsurgically managed patients with severe aortic stenosis. *Ann Thorac Surg* 2006;**82**:2111–2115.
6. Bramstedt KA. Aortic valve replacement in the elderly: frequently indicated yet frequently denied. *Gerontology* 2003;**49**:46–49.

7. Iung B, Cachier A, Baron G, Messika-Zeitoun D, Delahaye F, Tornos P, Gohlke-Barwolf C, Boersma E, Ravaut P, Vahanian A. Decision-making in elderly patients with severe aortic stenosis: why are so many denied surgery? *Eur Heart J* 2005;**26**:2714–2720.
8. Piazza N, Yoshinobu O, de Jaegere P, Serruys PW. Guidelines for reporting cardiac mortality and morbidity after cardiac valve interventions—need for a reappraisal? *Ann Thorac Surg* 2009;**87**:357–358.
9. Akins CW, Blackstone EH, Miller C, Kouchoukos NT, Turina MI. Guidelines for reporting cardiac mortality and morbidity after cardiac valve interventions—need for a reappraisal? (Response) *Ann Thorac Surg* 2009;**87**:359–360.
10. Thomas M, Wendler O. Transcatheter aortic valve implantation (TAVI): how to interpret the data and what data is required? *EuroIntervention* 2008;**5**:25–27.
11. Piazza N, Cutlip DE, Yoshinobu O, Kappetein AP, de Jaegere P, Serruys PW. Clinical endpoints in transcatheter aortic valve implantation: a call to ARC for standardised definitions. *EuroIntervention* 2008;**5**:28–29.
12. Akins CW, Miller DC, Turina MI et al. Guidelines for reporting mortality and morbidity after cardiac valve interventions. *J Thorac Cardiovasc Surg* 2008;**135**:732–738.
13. Thygesen K, Alpert JS, White HD, Jaffe AS, Apple FS, Galvani M, Katus HA, Newby LK, Ravkilde J, Chaitman B, Clemmensen PM, Dellborg M, Hod H, Porela P, Underwood R, Bax JJ, Beller GA, Bonow R, Van der Wall EE, Bassand JP, Wijns W, Ferguson TB, Steg PG, Uretsky BF, Williams DO, Armstrong PW, Antman EM, Fox KA, Hamm CW, Ohman EM, Simoons ML, Poole-Wilson PA, Gurfinkel EP, Lopez-Sendon JL, Pais P, Mendis S, Zhu JR, Wallentin LC, Fernandez-Aviles F, Fox KM, Parkhomenko AN, Priori SG, Tendera M, Voipio-Pulkki LM, Vahanian A, Camm AJ, De Caterina R, Dean V, Dickstein K, Filippatos G, Funck-Brentano C, Hellemans I, Kristensen SD, McGregor K, Sechtem U, Silber S, Widimsky P, Zamorano JL, Morais J, Brener S, Harrington R, Morrow D, Lim M, Martinez-Rios MA, Steinhilb S, Levine GN, Gibler WB, Goff D, Tubaro M, Dudek D, Al-Attar N. Universal definition of myocardial infarction. *J Am Coll Cardiol* 2007;**50**:2173–2195.
14. Cutlip DE, Windecker S, Mehran R, Boam A, Cohen DJ, van Es GA, Steg PG, Morel MA, Mauri L, Vranckx P, McFadden E, Lansky A, Hamon M, Krucoff MW, Serruys PW. Clinical end points in coronary stent trials: a case for standardized definitions. *Circulation* 2007;**115**:2344–2351.
15. Antonini-Canterin F, Leiballi E, Capanna M, Burelli C, Cassin M, Macor F, Grandis U, Nicolosi GL. Association between carotid and coronary artery disease in patients with aortic valve stenosis: an angiographic study. *Angiology* 2009;**60**:596–600.
16. Kahlert P, Knipp SC, Schlamann M, Theilmann M, Al-Rashid F, Weber M, Johansson U, Wendt D, Jakob HG, Forsting M, Sack S, Erbel R, Eggebrecht H. Silent and apparent cerebral ischemia after percutaneous transfemoral aortic valve implantation: a diffusion-weighted magnetic resonance imaging study. *Circulation* 2010;**121**:870–878.
17. Ghanem A, Muller A, Nahle CP, Kocurek J, Werner N, Hammerstingl C, Schild HH, Schwab JO, Mellert F, Fimmers R, Nickenig G, Thomas D. Risk and fate of cerebral embolism after transfemoral aortic valve implantation: a prospective pilot study with diffusion-weighted magnetic resonance imaging. *J Am Coll Cardiol* 2010;**55**:1427–1432.
18. Easton JD, Saver JL, Albers GW, Alberts MJ, Chaturvedi S, Feldmann E, Hatsukami TS, Higashida RT, Johnston SC, Kidwell CS, Lutsep HL, Miller E, Sacco RL. Definition and evaluation of transient ischemic attack: a scientific statement for healthcare professionals from the American Heart Association/American Stroke Association Stroke Council; Council on Cardiovascular Surgery and Anesthesia; Council on Cardiovascular Radiology and Intervention; Council on Cardiovascular Nursing; and the Interdisciplinary Council on Peripheral Vascular Disease. The American Academy of Neurology affirms the value of this statement as an educational tool for neurologists. *Stroke* 2009;**40**:2276–2293.
19. Saver JL. Proposal for a universal definition of cerebral infarction. *Stroke* 2008;**39**:3110–3115.
20. Ong AT, Serruys PW, Mohr FW, Morice MC, Kappetein AP, Holmes DR Jr, Mack MJ, van den Brand M, Morel MA, van Es GA, Kleijne J, Koglin J, Russell ME. The SYNergy between percutaneous coronary intervention with TAXus and cardiac surgery (SYNTAX) study: design, rationale, and run-in phase. *Am Heart J* 2006;**151**:1194–1204.
21. Farkouh ME, Dangas G, Leon MB, Smith C, Nesto R, Buse JB, Cohen DJ, Mahoney E, Sleeper L, King S 3rd, Domanski M, McKinlay S, Fuster V. Design of the Future REvascularization Evaluation in patients with Diabetes mellitus: Optimal management of Multivessel disease (FREEDOM) trial. *Am Heart J* 2008;**155**:215–223.
22. Gurm HS, Yadav JS, Fayad P, Katzen BT, Mishkel GJ, Bajwa TK, Ansel G, Strickman NE, Wang H, Cohen SA, Massaro JM, Cutlip DE. Long-term results of carotid stenting versus endarterectomy in high-risk patients. *N Engl J Med* 2008;**358**:1572–1579.
23. Ringleb PA, Allenberg J, Bruckmann H, Eckstein HH, Fraedrich G, Hartmann M, Hennerici M, Jansen O, Klein G, Kunze A, Marx P, Niederkorn K, Schmiedt W, Solymsi L, Stinglele R, Zeumer H, Hacke W. 30 day results from the SPACE trial of stent-protected angioplasty versus carotid endarterectomy in symptomatic patients: a randomised non-inferiority trial. *Lancet* 2006;**368**:1239–1247.
24. Hobson RW II, Howard VJ, Roubin GS, Brott TG, Ferguson RD, Popma JJ, Graham DL, Howard G. Carotid artery stenting is associated with increased complications in octogenarians: 30-day stroke and death rates in the CREST lead-in phase. *J Vasc Med Biol* 2004;**16**:1106–1111.
25. Maggioni AP, Franzosi MG, Santoro E, White H, Van de Werf F, Tognoni G. The risk of stroke in patients with acute myocardial infarction after thrombolytic and antithrombotic treatment. Gruppo Italiano per lo Studio della Sopravvivenza nell'Infarto Miocardico II (GISSI-2), and The International Study Group. *N Engl J Med* 1992;**327**:1–6.
26. ISIS-3 (Third International Study of Infarct Survival) Collaborative Group. ISIS-3: a randomised comparison of streptokinase vs tissue plasminogen activator vs anistreplase and of aspirin plus heparin vs aspirin alone among 41,299 cases of suspected acute myocardial infarction. *Lancet* 1992;**339**:753–770.
27. Gore JM, Granger CB, Simoons ML, Sloan MA, Weaver WD, White HD, Barbash GI, Van de Werf F, Aylward PE, Topol EJ. Stroke after thrombolysis. Mortality and functional outcomes in the GUSTO-I trial. Global use of strategies to open occluded coronary arteries. *Circulation* 1995;**92**:2811–2818.
28. Halperin JL. Ximelagatran compared with warfarin for prevention of thromboembolism in patients with nonvalvular atrial fibrillation: Rationale, objectives, and design of a pair of clinical studies and baseline patient characteristics (SPORTIF III and V). *Am Heart J* 2003;**146**:431–438.
29. Ezekowitz MD, Connolly S, Parekh A, Reilly PA, Varrone J, Wang S, Oldgren J, Themeles E, Wallentin L, Yusuf S. Rationale and design of RE-LY: randomized evaluation of long-term anticoagulant therapy, warfarin, compared with dabigatran. *Am Heart J* 2009;**157**:805–810, 810 e1–2.
30. Totaro P, Toni D, Durastanti L, Bozzao L, Gualdi GF, Raz E, Kouleridou A, Pantano P. Diffusion-weighted MRI in patients with non-diagnostic CT in the post-acute phase of cerebral ischemia. *Eur Neurol* 2010;**63**:94–100.
31. Brott T, Adams HP Jr, Olinger CP, Adams HP Jr, Tomsick T, Barsan WG, Biller J, Eberle R, Hertzberg V, Walker M. Measurements of acute cerebral infarction: a clinical examination scale. *Stroke* 1989;**20**:864–870.
32. Bonita R, Beaglehole R. Modification of Rankin Scale: recovery of motor function after stroke. *Stroke* 1988;**19**:1497–1500.
33. Goldstein LB, Samsa GP. Reliability of the National Institutes of Health Stroke Scale. Extension to non-neurologists in the context of a clinical trial. *Stroke* 1997;**28**:307–310.
34. Lyden PD, Lau GT. A critical appraisal of stroke evaluation and rating scales. *Stroke* 1991;**22**:1345–1352.
35. Standardized Definitions for Cardiovascular Outcomes Trials: FDA Draft Recommendations (personal communication; 24 March 2010).
36. Manoukian SV, Voeltz MD, Eikelboom J. Bleeding complications in acute coronary syndromes and percutaneous coronary intervention: predictors, prognostic significance, and paradigms for reducing risk. *Clin Cardiol* 2007;**30**:1124–1134.
37. Feit F, Voeltz MD, Attubato MJ, Lincoff AM, Chew DP, Bittl JA, Topol EJ, Manoukian SV. Predictors and impact of major hemorrhage on mortality following percutaneous coronary intervention from the REPLACE-2 trial. *Am J Cardiol* 2007;**100**:1364–1369.
38. Ndrepepa G, Berger PB, Mehilli J, Seyfarth M, Neumann FJ, Schömig A, Kastrati A. Periprocedural bleeding and 1-year outcome after percutaneous coronary interventions: appropriateness of including bleeding as a component of a quadruple end point. *J Am Coll Cardiol* 2008;**51**:690–697.
39. Budaj A, Eikelboom JW, Mehta SR, Afzal R, Chrolavicius S, Bassand JP, Fox KA, Wallentin L, Peters RJ, Granger CB, Joyner CD, Yusuf S; OASIS 5 Investigators. Improving clinical outcomes by reducing bleeding in patients with non-ST-elevation acute coronary syndromes. *Eur Heart J* 2009;**30**:655–661.
40. Mehran R, Pocock SJ, Stone GW, Clayton TC, Dangas GD, Feit F, Manoukian SV, Nikolsky E, Lansky AJ, Kirtane A, White HD, Colombo A, Ware JH, Moses JW, Ohman EM. Associations of major bleeding and myocardial infarction with the incidence and timing of mortality in patients presenting with non-ST-elevation acute coronary syndromes: a risk model from the ACUITY trial. *Eur Heart J* 2009;**30**:1457–1466.
41. Rao SV, Jollis JG, Harrington RA, Granger CB, Newby LK, Armstrong PW, Moliterno DJ, Lindblad L, Pieper K, Topol EJ, Stamler JS, Califf RM. Relationship of blood transfusion and clinical outcomes in patients with acute coronary syndromes. *JAMA* 2004;**292**:1555–1562.
42. Yang X, Alexander KP, Chen AY, Roe MT, Brindis RG, Rao SV, Gibler WB, Ohman EM, Peterson ED. The implications of blood transfusions for patients with non-ST-segment elevation acute coronary syndromes: results from the CRUSADE National Quality Improvement Initiative. *J Am Coll Cardiol* 2005;**46**:1490–1495.

43. Nikolsky E, Mehran R, Sadeghi HM, Grines CL, Cox DA, Garcia E, Tchong JE, Griffin JJ, Guagliumi G, Stuckey T, Turco M, Fahy M, Lansky AJ, Stone GW. Prognostic impact of blood transfusion after primary angioplasty for acute myocardial infarction: analysis from the CADILLAC (Controlled Abciximab and Device Investigation to Lower Late Angioplasty Complications) trial. *JACC Cardiovasc Interv* 2009;**2**:624–632.
44. Shishehbor MH, Madhwal S, Rajagopal V, Hsu A, Kelly P, Gurm HS, Kapadia SR, Lauer MS, Topol EJ. Impact of blood transfusion on short- and long-term mortality in patients with ST-segment elevation myocardial infarction. *JACC Cardiovasc Interv* 2009;**2**:46–53.
45. The GUSTO investigators. An international randomized trial comparing four thrombolytic strategies for acute myocardial infarction. *N Engl J Med* 1993;**329**:673–682.
46. Assessment of the Safety and Efficacy of a New Thrombolytic Regimen (ASSENT)-3 Investigators. Efficacy and safety of tenecteplase in combination with enoxaparin, abciximab, or unfractionated heparin: the ASSENT-3 randomized trial in acute myocardial infarction. *Lancet* 2001;**358**:605–613.
47. White H. Thrombin-specific anticoagulation with bivalirudin versus heparin in patients receiving fibrinolytic therapy for acute myocardial infarction: the HERO-2 randomised trial. *Lancet* 2001;**358**:1855–1863.
48. Mehran R, Brodie B, Cox DA, Grines CL, Rutherford B, Bhatt DL, Dangas G, Feit F, Ohman EM, Parise H, Fahy M, Lansky AJ, Stone GW. The Harmonizing Outcomes with Revascularization Stents in Acute Myocardial Infarction (HORIZONS-AMI) trial: study design and rationale. *Am Heart J* 2008;**156**:44–56.
49. Antman EM, McCabe CH, Gurfinkel EP, Turpie AG, Bernink PJ, Salein D, Bayes De Luna A, Fox K, Lablanche JM, Radley D, Premmureur J, Braunwald E. Enoxaparin prevents death and cardiac ischemic events in unstable anginal/non-Q-wave myocardial infarction. Results of the thrombolysis in myocardial infarction (TIMI) 11B trial. *Circulation* 1999;**100**:1593–1601.
50. Cohen M, Demers C, Gurfinkel EP, Turpie AG, Fromell GJ, Goodman S, Langer A, Califf RM, Fox KA, Premmureur J, Bigonzi F. A comparison of low-molecular-weight heparin with unfractionated heparin for unstable coronary artery disease. Efficacy and Safety of Subcutaneous Enoxaparin in Non-Q-Wave Coronary Events Study Group. *N Engl J Med* 1997;**337**:447–452.
51. Goodman SG, Fitchett D, Armstrong PW, Tan M, Langer A. Randomized evaluation of the safety and efficacy of enoxaparin versus unfractionated heparin in high-risk patients with non-ST-segment elevation acute coronary syndromes receiving the glycoprotein IIb/IIIa inhibitor eptifibatid. *Circulation* 2003;**107**:238–244.
52. Blazing MA, de Lemos JA, White HD, Fox KA, Verheugt FW, Ardissino D, DiBattiste PM, Palmisano J, Bilheimer DW, Snapinn SM, Ramsey KE, Gardner LH, Hasselblad V, Pfeffer MA, Lewis EF, Braunwald E, Califf RM. Safety and efficacy of enoxaparin vs unfractionated heparin in patients with non-ST-segment elevation acute coronary syndromes who receive tirofiban and aspirin: a randomized controlled trial. *JAMA* 2004;**292**:55–64.
53. SYNERGY Executive Committee. The SYNERGY trial: study design and rationale. *Am Heart J* 2002;**143**:952–960.
54. Yusuf S, Mehta SR, Chrolavicius S, Afzal R, Pogue J, Granger CB, Budaj A, Peters RJ, Bassand JP, Wallentin L, Joyner C, Fox KA. Comparison of fondaparinux and enoxaparin in acute coronary syndromes. *N Engl J Med* 2006;**354**:1464–1476.
55. Lincoff AM, Kleiman NS, Kereiakes DJ, Feit F, Bittl JA, Jackman JD, Sarembok IJ, Cohen DJ, Spriggs D, Ebrahimi R, Keren G, Carr J, Cohen EA, Betriu A, Desmet W, Rutsch W, Wilcox RG, de Feyter PJ, Vahanian A, Topol EJ. Long-term efficacy of bivalirudin and provisional glycoprotein IIb/IIIa blockade vs heparin and planned glycoprotein IIb/IIIa blockade during percutaneous coronary revascularization: REPLACE-2 randomized trial. *JAMA* 2004;**292**:696–703.
56. Stone GW, Bertrand M, Colombo A, Dangas G, Farkouh ME, Feit F, Lansky AJ, Lincoff AM, Mehran R, Moses JW, Ohman M, White HD. Acute Catheterization and Urgent Intervention Triage strategy (ACUITY) trial: study design and rationale. *Am Heart J* 2004;**148**:764–775.
57. Subherwal S, Bach RG, Chen AY, Gage BF, Rao SV, Newby LK, Wang TY, Gibler WB, Ohman EM, Roe MT, Pollack CV Jr, Peterson ED, Alexander KP. Baseline risk of major bleeding in non-ST-segment-elevation myocardial infarction: the CRUSADE (Can Rapid risk stratification of Unstable angina patients Suppress ADverse outcomes with Early implementation of the ACC/AHA Guidelines) bleeding score. *Circulation* 2009;**119**:1873–1882.
58. Mehta SR, Bassand JP, Chrolavicius S, Diaz R, Fox KA, Granger CB, Jolly S, Rupprecht HJ, Widimsky P, Yusuf S. Design and rationale of CURRENT-OASIS 7: a randomized, 2 × 2 factorial trial evaluating optimal dosing strategies for clopidogrel and aspirin in patients with ST and non-ST-elevation acute coronary syndromes managed with an early invasive strategy. *Am Heart J* 2008;**156**:1080–1088 e1.
59. James S, Akerblom A, Cannon CP, Emanuelsson H, Husted S, Katus H, Skene A, Steg PG, Storey RF, Harrington R, Becker R, Wallentin L. Comparison of ticagrelor, the first reversible oral P2Y<sub>12</sub> receptor antagonist, with clopidogrel in patients with acute coronary syndromes: rationale, design, and baseline characteristics of the PLATelet inhibition and patient Outcomes (PLATO) trial. *Am Heart J* 2009;**157**:599–605.
60. Schulman S, Kearon C. Definition of major bleeding in clinical investigations of antihemostatic medicinal products in non-surgical patients. *J Thromb Haemost* 2005;**3**:692–694.
61. Levy EM, Viscoli CM, Horwitz RI. The effect of acute renal failure on mortality. A cohort analysis. *JAMA* 1996;**275**:1489–1494.
62. McCullough PA, Wolyn R, Rocher LL, Levin RN, O'Neill VWW. Acute renal failure after coronary intervention: incidence, risk factors, and relationship to mortality. *Am J Med* 1997;**103**:368–375.
63. Mangano CM, Diamondstone LS, Ramsay JG, Aggarwal A, Herskowitz A, Mangano DT. Renal dysfunction after myocardial revascularization: risk factors, adverse outcomes, and hospital resource utilization. The Multicenter Study of Perioperative Ischemia Research Group. *Ann Intern Med* 1998;**128**:194–203.
64. Gruberg L, Mintz GS, Mehran R, Dangas G, Lansky AJ, Kent KM, Pichard AD, Satler LF, Leon MB. The prognostic implications of further renal function deterioration within 48 h of interventional coronary procedures in patients with pre-existent chronic renal insufficiency. *J Am Coll Cardiol* 2000;**36**:1542–1548.
65. Best PJ, Lennon R, Ting HH, Bell MR, Rihal CS, Holmes DR, Berger PB. The impact of renal insufficiency on clinical outcomes in patients undergoing percutaneous coronary interventions. *J Am Coll Cardiol* 2002;**39**:1113–1119.
66. Nash K, Hafeez A, Hou S. Hospital-acquired renal insufficiency. *Am J Kidney Dis* 2002;**39**:930–936.
67. Rihal CS, Textor SC, Grill DE, Berger PB, Ting HH, Best PJ, Singh M, Bell MR, Barsness GW, Mathew V, Garratt KN, Holmes DR Jr. Incidence and prognostic importance of acute renal failure after percutaneous coronary intervention. *Circulation* 2002;**105**:2259–2264.
68. Antunes PE, Prieto D, Ferrao de Oliveira J, Antunes MJ. Renal dysfunction after myocardial revascularization. *Eur J Cardiothorac Surg* 2004;**25**:597–604.
69. Lassnigg A, Schmidlin D, Mouhieddine M, Bachmann LM, Druml W, Bauer P, Hiesmayr M. Minimal changes of serum creatinine predict prognosis in patients after cardiothoracic surgery: a prospective cohort study. *J Am Soc Nephrol* 2004;**15**:1597–1605.
70. Chertow GM, Burdick E, Honour M, Bonventre JV, Bates DW. Acute kidney injury, mortality, length of stay, and costs in hospitalized patients. *J Am Soc Nephrol* 2005;**16**:3365–3370.
71. Aregger F, Wenaweser P, Hellige GJ, Kadner A, Carrel T, Windecker S, Frey FJ. Risk of acute kidney injury in patients with severe aortic valve stenosis undergoing transcatheter valve replacement. *Nephrol Dial Transplant* 2009;**24**:2175–2179.
72. Bagur R, Webb JG, Nietlispach F, Dumont E, De Larochelliere R, Doyle D, Masson JB, Gutierrez MJ, Clavel MA, Bertrand OF, Pibarot P, Rodes-Cabau J. Acute kidney injury following transcatheter aortic valve implantation: predictive factors, prognostic value, and comparison with surgical aortic valve replacement. *Eur Heart J* 2010;**31**:865–874.
73. Lameire N, Van Biesen W, Vanholder R. The changing epidemiology of acute renal failure. *Nat Clin Pract Nephrol* 2006;**2**:364–377.
74. Bellomo R, Ronco C, Kellum JA, Mehta RL, Palevsky P. Acute renal failure definition, outcome measures, animal models, fluid therapy and information technology needs: the Second International Consensus Conference of the Acute Dialysis Quality Initiative (ADQI) Group. *Crit Care* 2004;**8**:R204–R212.
75. Bellomo R, Kellum JA, Ronco C. Defining acute renal failure: physiological principles. *Intensive Care Med* 2004;**30**:33–37.
76. Kellum JA, Levin N, Bouman C, Lameire N. Developing a consensus classification system for acute renal failure. *Curr Opin Crit Care* 2002;**8**:509–514.
77. Abosaif NY, Tolba YA, Heap M, Russell J, El Nahas AM. The outcome of acute renal failure in the intensive care unit according to RIFLE: model application, sensitivity, and predictability. *Am J Kidney Dis* 2005;**46**:1038–1048.
78. Bell M, Liljestam E, Granath F, Fryckstedt J, Ekblom A, Martling CR. Optimal follow-up time after continuous renal replacement therapy in actual renal failure patients stratified with the RIFLE criteria. *Nephrol Dial Transplant* 2005;**20**:354–360.
79. Hoste EA, Clermont G, Kersten A, Venkataraman R, Angus DC, De Bacquer D, Kellum JA. RIFLE criteria for acute kidney injury are associated with hospital mortality in critically ill patients: a cohort analysis. *Crit Care* 2006;**10**:R73.
80. Kuitunen A, Vento A, Suojaranta-Ylinen R, Pettila V. Acute renal failure after cardiac surgery: evaluation of the RIFLE classification. *Ann Thorac Surg* 2006;**81**:542–546.
81. Uchino S, Bellomo R, Goldsmith D, Bates S, Ronco C. An assessment of the RIFLE criteria for acute renal failure in hospitalized patients. *Crit Care Med* 2006;**34**:1913–1917.

82. Mehta RL, Kellum JA, Shah SV, Molitoris BA, Ronco C, Warnock DG, Levin A. Acute kidney injury network: report of an initiative to improve outcomes in acute kidney injury. *Crit Care* 2007;**11**:R31.
83. Guierrez NV, Diaz A, Timmis GC, O'Neill WW, Stevens MA, Sandberg KR, McCullough PA. Determinants of serum creatinine trajectory in acute contrast nephropathy. *J Interv Cardiol* 2002;**15**:349–354.
84. Van Mieghem NM, Nuis RJ, Piazza N, Apostolos T, Ligthart J, Schultz C, de Jaegere PP, Serruys PW. Vascular complications with transcatheter aortic valve implantation using the 18 Fr Medtronic CoreValve System(R): the Rotterdam experience. *EuroIntervention* 2010;**5**:673–679.
85. Ducrocq G, Francis F, Serfaty JM, Himbert D, Maury JM, Pasi N, Marouene S, Provenchere S, lung B, Castier Y, Leseche G, Vahanian A. Vascular complications of transfemoral aortic valve implantation with the Edwards SAPIEN prosthesis: incidence and impact on outcome. *EuroIntervention* 2010;**5**:666–672.
86. Tchetché D, Dumonteil N, Sauguet A, Descoutures F, Luz A, Garcia O, Soula P, Gabiache Y, Fournial G, Marcheix B, Carrie D, Fajadet J. Thirty-day outcome and vascular complications after transarterial aortic valve implantation using both Edwards Sapien and Medtronic CoreValve(R) bioprostheses in a mixed population. *EuroIntervention* 2010;**5**:659–665.
87. Kahlert P, Al-Rashid F, Weber M, Wendt D, Heine T, Kottenberg E, Thielmann M, Kuhl H, Peters J, Jakob HG, Sack S, Erbel R, Eggebrecht H. Vascular access site complications after percutaneous transfemoral aortic valve implantation. *Herz* 2009;**34**:398–408.
88. Chaikof EL, Blankensteijn JD, Harris PL, White GH, Zarins CK, Bernhard VM, Matsumura JS, May J, Veith FJ, Fillinger MF, Rutherford RB, Kent KC. Reporting standards for endovascular aortic aneurysm repair. *J Vasc Surg* 2002;**35**:1048–1060.
89. Bleiziffer S, Ruge H, Mazzitelli D, Hutter A, Opitz A, Bauernschmitt R, Lange R. Survival after transapical and transfemoral aortic valve implantation: talking about two different patient populations. *J Thorac Cardiovasc Surg* 2009;**138**:1073–1080.
90. Zoghbi WA, Chambers JB, Dumesnil JG, Foster E, Gottdiener JS, Grayburn PA, Khandheria BK, Levine RA, Marx GR, Miller FA Jr., Nakatani S, Quinones MA, Rakowski H, Rodriguez LL, Swaminathan M, Waggoner AD, Weissman NJ, Zabalgoitia M. Recommendations for evaluation of prosthetic valves with echocardiography and doppler ultrasound: a report From the American Society of Echocardiography's Guidelines and Standards Committee and the Task Force on Prosthetic Valves, developed in conjunction with the American College of Cardiology Cardiovascular Imaging Committee, Cardiac Imaging Committee of the American Heart Association, the European Association of Echocardiography, a registered branch of the European Society of Cardiology, the Japanese Society of Echocardiography and the Canadian Society of Echocardiography, endorsed by the American College of Cardiology Foundation, American Heart Association, European Association of Echocardiography, a registered branch of the European Society of Cardiology, the Japanese Society of Echocardiography, and Canadian Society of Echocardiography. *J Am Soc Echocardiogr* 2009;**22**:975–1014; quiz 1082–4.
91. Vahanian A, Baumgartner H, Bax J, Butchart E, Dion R, Filippatos G, Flachskampf F, Hall R, lung B, Kasprzak J, Nataf P, Tornos P, Torracca L, Wenink A. Guidelines on the management of valvular heart disease: The Task Force on the Management of Valvular Heart Disease of the European Society of Cardiology. *Eur Heart J* 2007;**28**:230–268.
92. Bonow RO, Carabello BA, Chatterjee K, de Leon AC Jr, Faxon DP, Freed MD, Gaasch WH, Lytle BW, Nishimura RA, O'Gara PT, O'Rourke RA, Otto CM, Shah PM, Shanewise JS. Focused update incorporated into the ACC/AHA 2006 guidelines for the management of patients with valvular heart disease: a report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines (Writing Committee to Revise the 1998 Guidelines for the Management of Patients With Valvular Heart Disease): endorsed by the Society of Cardiovascular Anesthesiologists, Society for Cardiovascular Angiography and Interventions, and Society of Thoracic Surgeons. *Circulation* 2008;**118**:e523–e661.
93. Schultz CJ, Weustink A, Piazza N, Otten A, Mollet N, Krestin G, van Geuns RJ, de Feyter P, Serruys PW, de Jaegere P. Geometry and degree of apposition of the CoreValve ReValving system with multislice computed tomography after implantation in patients with aortic stenosis. *J Am Coll Cardiol* 2009;**54**:911–918.
94. Schultz C, Piazza N, van Geuns RJ, de Feyter P, Serruys P, de Jaegere P. *Tips and Tricks for Imaging Percutaneously Implanted Aortic Valve Prostheses*. New York: Informa Healthcare, 2010. p275–287.
95. Cribier A, Eltchaninoff H, Tron C, Bauer F, Agatiello C, Nercolini D, Tapiero S, Litzler PY, Bessou JP, Babaliaros V. Treatment of calcific aortic stenosis with the percutaneous heart valve: mid-term follow-up from the initial feasibility studies: the French experience. *J Am Coll Cardiol* 2006;**47**:1214–1223.
96. Webb JG, Altwegg L, Boone RH, Cheung A, Ye J, Lichtenstein S, Lee M, Masson JB, Thompson C, Moss R, Carere R, Munt B, Nietlisbach F, Humphries K. Transcatheter aortic valve implantation: impact on clinical and valve-related outcomes. *Circulation* 2009;**119**:3009–3016.
97. Eberhard Grube LB, Ralf M, Barthel S, Bernfried Z, Dinesh N, Harald B, Thomas F, Stein I, Ulrich G. Progress and current status of percutaneous aortic valve replacement: results of three device generations of the corevalve revalving system. *Circ Cardiovasc Interv* 2008;**1**:167–175.
98. Piazza N, Grube E, Gerckens U, den Heijer P, Linke A, Luha O, Ramondo A, Ussia G, Wenaweser P, Windecker S, Laborde JC, de Jaegere P, Serruys PW. Procedural and 30-day outcomes following transcatheter aortic valve implantation using the third generation (18 Fr) corevalve revalving system: results from the multicentre, expanded evaluation registry 1-year following CE mark approval. *EuroIntervention* 2008;**4**:242–249.
99. Walther T, Simon P, Dewey T, Wimmer-Greinecker G, Falk V, Kasimir MT, Doss M, Borger MA, Schuler G, Glogar D, Fehske W, Wolner E, Mohr FW, Mack M. Transapical minimally invasive aortic valve implantation: multicenter experience. *Circulation* 2007;**116**:1240–1245.
100. Svensson LG, Dewey T, Kapadia S, Roselli EE, Stewart A, Williams M, Anderson WN, Brown D, Leon M, Lytle B, Moses J, Mack M, Tuzcu M, Smith C. United States feasibility study of transcatheter insertion of a stented aortic valve by the left ventricular apex. *Ann Thorac Surg* 2008;**86**:46–54.
101. Himbert D, Descoutures F, Al-Attar N, lung B, Ducrocq G, Détaint D, Brochet E, Messika-Zeitoun D, Francis F, Ibrahim H, Nataf P, Vahanian A. Results of transfemoral or transcatheter aortic valve implantation following a uniform assessment in high-risk patients with aortic stenosis. *J Am Coll Cardiol* 2009;**54**:303–311.
102. De Jaegere P, Piazza N, Galema T, Otten A, Soliman O, Van Dalen B, Geleijnse M, Kappetein A, Garcia Garcia H, Van Es G, Serruys P. Early echocardiographic evaluation following percutaneous implantation with the self-expanding CoreValve ReValving System aortic valve bioprosthesis. *Eurointervention* 2008;**4**:351–357.
103. Détaint D, Lepage L, Himbert D, Brochet E, Messika-Zeitoun D, lung B, Vahanian A. Determinants of significant paravalvular regurgitation after transcatheter aortic valve: implantation impact of device and annulus discongruence. *JACC Cardiovasc Interv* 2009;**2**:821–827.
104. Clavel MA, Webb JG, Pibarot P, Altwegg L, Dumont E, Thompson C, De Larochellière R, Doyle D, Masson JB, Bergeron S, Bertrand OF, Rodés-Cabau J. Comparison of the hemodynamic performance of percutaneous and surgical bioprostheses for the treatment of severe aortic stenosis. *J Am Coll Cardiol* 2009;**53**:1883–1891.
105. Davila-Roman VG, Waggoner AD, Kennard ED, Holubkov R, Jamieson WR, Englberger L, Carrel TP, Schaff HV. Prevalence and severity of paravalvular regurgitation in the Artificial Valve Endocarditis Reduction Trial (AVERT) echocardiography study. *J Am Coll Cardiol* 2004;**44**:1467–1472.
106. Ionescu A, Fraser AG, Butchart EG. Prevalence and clinical significance of incidental paraprosthetic valvar regurgitation: a prospective study using transoesophageal echocardiography. *Heart* 2003;**89**:1316–1321.
107. O'Rourke DJ, Palac RT, Malenka DJ, Marrin CA, Arbuuckle BE, Plehn JF. Outcome of mild paraprosthetic regurgitation detected by intraoperative transesophageal echocardiography. *J Am Coll Cardiol* 2001;**38**:163–166.
108. Rallidis LS, Moysakakis IE, Ikonomidis I, Nihoyannopoulos P. Natural history of early aortic paraprosthetic regurgitation: a five-year follow-up. *Am Heart J* 1999;**138**:351–357.
109. Kupferwasser II, Mohr-Kahaly S, Erbel R, Nolting A, Dohmen G, Oelert H, Meyer J. Improved assessment of pathological regurgitation in patients with prosthetic heart valves by multiplane transesophageal echocardiography. *Echocardiography* 1997;**14**:363–374.
110. Mohr-Kahaly S, Kupferwasser I, Erbel R, Oelert H, Meyer J. Regurgitant flow in apparently normal valve prostheses: improved detection and semiquantitative analysis by transesophageal two-dimensional color-coded Doppler echocardiography. *J Am Soc Echocardiogr* 1990;**3**:187–195.
111. Mohr-Kahaly S, Kupferwasser I, Erbel R, Wittlich N, Iversen S, Oelert H, Meyer J. Value and limitations of transesophageal echocardiography in the evaluation of aortic prostheses. *J Am Soc Echocardiogr* 1993;**6**:12–20.
112. Trepels T, Martens S, Doss M, Fichtlscherer S, Schächinger V. Images in cardiovascular medicine. Thrombotic restenosis after minimally invasive implantation of aortic valve stent. *Circulation* 2009;**120**:e23–e24.
113. Durack DT, Lukes AS, Bright DK. New criteria for diagnosis of infective endocarditis: utilization of specific echocardiographic findings. Duke Endocarditis Service. *Am J Med* 1994;**96**:200–209.
114. Comoglio C, Boffini M, El Qarra S, Sansone F, D'Amico M, Marra S, Rinaldi M. Aortic valve replacement and mitral valve repair as treatment of complications after percutaneous core valve implantation. *J Thorac Cardiovasc Surg* 2009;**138**:1025–1027.
115. Wong DR, Boone RH, Thompson CR, Allard MF, Altwegg L, Carere RG, Cheung A, Ye J, Lichtenstein SV, Ling H, Webb JG. Mitral valve injury late

- after transcatheter aortic valve implantation. *J Thorac Cardiovasc Surg* 2009;**137**: 1547–1549.
116. Piazza N, Marra S, Webb J, D'Amico M, Rinaldi M, Boffini M, Comoglio C, Scacciarella P, Kappetein A, de Jaegere P, Serruys P. Two cases of Aneurysm of the Anterior Mitral Valve Leaflet associated with Transcatheter Aortic Valve Endocarditis (TAVE): a mere coincidence? *J Cardiothorac Surg* 2010;**140**: e36–8. Published online ahead of print 18 February 2010.
  117. Piazza N, de Jaegere P, Schultz C, Becker AE, Serruys PW, Anderson RH. Anatomy of the aortic valvar complex and its implications for transcatheter implantation of the aortic valve. *Circ Cardiovasc Intervent* 2008;**1**:74–81.
  118. Moreno R, Dobarro D, Lopez de Sa E, Prieto M, Morales C, Calvo Orbe L, Moreno-Gomez I, Filgueiras D, Sanchez-Recalde A, Galeote G, Jimenez-Valero S, Lopez-Sendon JL. Cause of complete atrioventricular block after percutaneous aortic valve implantation: insights from a necropsy study. *Circulation* 2009;**120**:e29–e30.
  119. El-Khalily ZTB, Staniloae C, Theroux P, Dubuc M, Roy D, Guerra P, Macle L, Talajic M. Prognostic significance of newly acquired bundle branch block after aortic valve replacement. *Am J Cardiol* 2004;**94**:1008–1011.
  120. Thomas JL, Dickstein RA, Parker FB Jr., Potts JL, Poirier RA, Fruehan CT, Eich RH. Prognostic significance of the development of left bundle conduction defects following aortic valve replacement. *J Thorac Cardiovasc Surg* 1982;**84**: 382–386.
  121. Limongelli G, Ducceschi V, D'Andrea A, Renzulli A, Sarubbi B, De Feo M, Cerasuolo F, Calabro R, Cotrufo M. Risk factors for pacemaker implantation following aortic valve replacement: a single centre experience. *Heart* 2003;**89**: 901–904.
  122. Gordon RS, Ivanov J, Cohen G, Ralph-Edwards AL. Permanent cardiac pacing after a cardiac operation: predicting the use of permanent pacemakers. *Ann Thorac Surg* 1998;**66**:1698–1704.
  123. Dawkins S, Hobson AR, Kalra PR, Tang AT, Monro JL, Dawkins KD. Permanent pacemaker implantation after isolated aortic valve replacement: incidence, indications, and predictors. *Ann Thorac Surg* 2008;**85**:108–112.
  124. Sinhal A, Altwegg L, Pasupati S, Allard M, Martin P, Cheung A, Ye J, Kerr C, Lichtenstein SV, Webb JG. Atrioventricular block after transcatheter balloon expandable aortic valve implantation. *JACC Cardiovasc Interv* 2008;**1**:305–309.
  125. Jilahiawi H, Chin D, Vasa-Nicotera M, Jeilan M, Spyt T, Ng GA, Bence J, Logtens E, Kovac J. Predictors for permanent pacemaker requirement after transcatheter aortic valve implantation with the CoreValve bioprosthesis. *Am Heart J* 2009;**157**:860–866.
  126. Calvi V, Puzangara E, Pruiti GP, Conti S, Di Grazia A, Ussia GP, Capodanno D, Tamburino C. Early conduction disorders following percutaneous aortic valve replacement. *Pacing Clin Electrophysiol* 2009;**32** Suppl 1:S126–S130.
  127. Gutierrez M, Rodes-Cabau J, Bagur R, Doyle D, DeLarochelliere R, Bergeron S, Lemieux J, Villeneuve J, Cote M, Bertrand OF, Poirier P, Clavel MA, Pibarot P, Dumont E. Electrocardiographic changes and clinical outcomes after transapical aortic valve implantation. *Am Heart J* 2009;**158**:302–308.
  128. Piazza N, Onuma Y, Jesserun E, Kint PP, Maugenest AM, Anderson RH, de Jaegere P, Serruys PW. Early and persistent intraventricular conduction abnormalities and requirements for pacemaking after percutaneous replacement of the aortic valve. *JACC Cardiovasc Interv* 2008;**1**:310–316.
  129. Orłowska-Baranowska E, Baranowski R, Michalek P, Hoffman P, Rywik T, Rawczyńska-Englert I. Prediction of paroxysmal atrial fibrillation after aortic valve replacement in patients with aortic stenosis: identification of potential risk factors. *J Heart Valve Dis* 2003;**12**:136–141.
  130. Banach M, Goch A, Misztal M, Rysz J, Jaszewski R, Goch JH. Predictors of paroxysmal atrial fibrillation in patients undergoing aortic valve replacement. *J Thorac Cardiovasc Surg* 2007;**134**:1569–1576.
  131. Webb JG, Pasupati S, Humphries K, Thompson C, Altwegg L, Moss R, Sinhal A, Carere RG, Munt B, Ricci D, Ye J, Cheung A, Lichtenstein SV. Percutaneous transarterial aortic valve replacement in selected high-risk patients with aortic stenosis. *Circulation* 2007;**116**:755–763.
  132. Webb JG. Coronary obstruction due to transcatheter valve implantation. *Catheter Cardiovasc Interv* 2009;**73**:973.
  133. Al-Attar N, Ghodbane W, Himbert D, Rau C, Raffoul R, Messika-Zeitoun D, Brochet E, Vahanian A, Nataf P. Unexpected complications of transapical aortic valve implantation. *Ann Thorac Surg* 2009;**88**:90–94.
  134. Tzikas A, Schultz C, Piazza N, van Geuns RJ, Serruys PW, de Jaegere PP. Perforation of the membranous interventricular septum after transcatheter aortic valve implantation. *Circ Cardiovasc Interv* 2009;**2**:582–583.
  135. Hunt SA, Abraham WT, Chin MH, Feldman AM, Francis GS, Ganiats TG, Jessup M, Konstam MA, Mancini DM, Michl K, Oates JA, Rahko PS, Silver MA, Stevenson LW, Yancy CW. Focused update incorporated into the ACC/AHA 2005 Guidelines for the Diagnosis and Management of Heart Failure in Adults: a report of the American College of Cardiology Foundation/American Heart Association Task Force on Practice Guidelines: developed in collaboration with the International Society for Heart and Lung Transplantation. *Circulation* 2009;**119**:e391–e479.
  136. Guyatt GH, Sullivan MJ, Thompson PJ, Fallen EL, Pugsley SO, Taylor DW, Berman LB. The 6-minute walk: a new measure of exercise capacity in patients with chronic heart failure. *Can Med Assoc J* 1985;**132**:919–923.
  137. Little B (ed.). *The Criteria Committee of the New York Heart Association. Diseases of the heart and blood vessels: nomenclature and criteria for diagnosis*. 6th ed. Boston: Little, Brown, 1964. Boston, 1964.
  138. Group TE. EuroQol—a new facility for the measurement of health-related quality of life. The EuroQol Group. *Health Policy* 1990;**16**:199–208.
  139. Fried LP, Tangen CM, Walston J, Newman AB, Hirsch C, Gottdiener J, Seeman T, Tracy R, Kop WJ, Burke G, McBurnie MA. Frailty in older adults: evidence for a phenotype. *J Gerontol A Biol Sci Med Sci* 2001;**56**:M146–M156.
  140. Green CP, Porter CB, Bresnahan DR, Spertus JA. Development and evaluation of the Kansas City Cardiomyopathy Questionnaire: a new health status measure for heart failure. *J Am Coll Cardiol* 2000;**35**:1245–1255.
  141. Brooks R, Rabin R, de Charro F (eds) *The Measurement and Valuation of Health Status Using EQ-5D: A European Perspective: Evidence from the EuroQol BIO MED Research Programme*. Rotterdam: Kluwer Academic Publishers; 2003. 2003.
  142. Spertus J, Peterson E, Conard MW, Heidenreich PA, Krumholz HM, Jones P, McCullough PA, Pina I, Tooley J, Weintraub WS, Rumsfeld JS. Monitoring clinical changes in patients with heart failure: a comparison of methods. *Am Heart J* 2005;**150**:707–715.
  143. Ussia GP, Mule M, Barbanti M, Cammalleri V, Scarabelli M, Imme S, Capodanno D, Ciriminna S, Tamburino C. Quality of life assessment after percutaneous aortic valve implantation. *Eur Heart J* 2009;**30**:1790–1796.
  144. Winograd CH, Gerety MB, Chung M, Goldstein MK, Dominguez F Jr, Vallone R. Screening for frailty: criteria and predictors of outcomes. *J Am Geriatr Soc* 1991;**39**:778–784.
  145. Hamerman D. Toward an understanding of frailty. *Ann Intern Med* 1999;**130**: 945–950.
  146. Lee DH, Butth KJ, Martin BJ et al. Frail patients are at increased risk for mortality and prolonged institutional care after cardiac surgery. *Circulation* 2010;**121**: 973–978.
  147. Disability and frailty among elderly Canadians: a comparison of six surveys. *Int Psychogeriatr* 2001;**13** Supp 1:159–167.
  148. Chin APMJ, de Groot LC, van Gend SV, Schoterman MH, Schouten EG, Schroll M, van Staveren WA. Inactivity and weight loss: effective criteria to identify frailty. *J Nutr Health Aging* 2003;**7**:55–60.
  149. Walther T, Falk V, Kempfert J, Borger MA, Fassl J, Chu MW, Schuler G, Mohr FW. Transapical minimally invasive aortic valve implantation: the initial 50 patients. *Eur J Cardiothorac Surg* 2008;**33**:983–988.
  150. Webb JG, Chandavimol M, Thompson CR, Ricci DR, Carere RG, Munt BI, Buller CE, Pasupati S, Lichtenstein S. Percutaneous aortic valve implantation retrograde from the femoral artery. *Circulation* 2006;**113**:842–850.
  151. Ferrari E, Rizzo E, Sulzer C, von Segesser LK. Unexpected left ventricular free-wall rupture following an aortic catheter-valve implantation. *Eur J Cardiothorac Surg* 2010;**37**:242–244. Published online ahead of print 19 August 2009.
  152. Maroto LC, Rodriguez JE, Cobiella J, Silva J. Delayed dislocation of a transapically implanted aortic bioprosthesis. *Eur J Cardiothorac Surg* 2009;**36**:935–937.
  153. Clavel MA, Dumont E, Pibarot P, Doyle D, De Larochelliere R, Villeneuve J, Bergeron S, Couture C, Rodes-Cabau J. Severe valvular regurgitation and late prosthesis embolization after percutaneous aortic valve implantation. *Ann Thorac Surg* 2009;**87**:621–623.
  154. Ruiz C, Laborde J, Condado J, Chiam P, Condado J. First percutaneous transcatheter aortic valve-in-valve implant with three year follow-up. *Catheter Cardiovasc Interv* 2008;**72**:143–148.
  155. Piazza N, Schultz C, de Jaegere PP, Serruys PW. Implantation of two self-expanding aortic bioprosthetic valves during the same procedure—Insights into valve-in-valve implantation ('Russian doll concept'). *Catheter Cardiovasc Interv* 2009;**73**:530–539.
  156. Montorsi P, De Bernardi F, Muratori M, Cavoretto D, Pepi M. Role of cine-fluoroscopy, transthoracic, and transesophageal echocardiography in patients with suspected prosthetic heart valve thrombosis. *Am J Cardiol* 2000;**85**:58–64.
  157. Cianciulli TE, Lax JA, Beck MA, Cerruti FE, Gigena GE, Saccheri MC, Fernandez E, Dorelle AN, Leguizamon JH, Prezioso HA. Cinefluoroscopic assessment of mechanical disc prostheses: its value as a complementary method to echocardiography. *J Heart Valve Dis* 2005;**14**:664–673.
  158. Piazza N, Grube E, Gerckens U, Schuler G, Linke A, den Heijer P, Kovacs J, Spyt T, Laborde J, Morel M, Nuis R, Garcia-Garcia H, de Jaegere P, Serruys P. A clinical protocol for analysis of the structural integrity of the Medtronic CoreValve System(R) frame and its application in patients with 1-year minimum follow-up. *EuroIntervention* 2010;**5**:680–686.

159. Thyregod HG, Lund JT, Engström T, Steinbruchel DA. Transcatheter aortic valve prosthesis surgically replaced 4 months after implantation. *Eur J Cardiothorac Surg* 2010;**37**:494–496.
160. Rosengart TK, Feldman T, Borger MA, Vassiliades TA Jr, Gillinov AM, Hoercher KJ, Vahanian A, Bonow RO, O'Neill W; American Heart Association Council on Cardiovascular Surgery and Anesthesia; American Heart Association Council on Clinical Cardiology; Functional Genomics and Translational Biology Interdisciplinary Working Group; Quality of Care and Outcomes Research Interdisciplinary Working Group. Percutaneous and minimally invasive valve procedures: a scientific statement from the American Heart Association Council on Cardiovascular Surgery and Anesthesia, Council on Clinical Cardiology, Functional Genomics and Translational Biology Interdisciplinary Working Group, and Quality of Care and Outcomes Research Interdisciplinary Working Group. *Circulation* 2008;**117**:1750–1767.
161. Vahanian A, Alfieri O, Al-Attar N, Antunes M, Bax J, Cormier B, Cribier A, De Jaegere P, Fournial G, Kappetein AP, Kovac J, Ludgate S, Maisano F, Moat N, Mohr F, Nataf P, Piérard L, Pomar JL, Schofer J, Tornos P, Tuzcu M, van Hout B, Von Segesser LK, Walther T; European Association of Cardio-Thoracic Surgery; European Society of Cardiology; European Association of Percutaneous Cardiovascular Interventions. Transcatheter valve implantation for patients with aortic stenosis: a position statement from the European Association of Cardio-Thoracic Surgery (EACTS) and the European Society of Cardiology (ESC), in collaboration with the European Association of Percutaneous Cardiovascular Interventions (EAPCI). *Eur Heart J* 2008;**29**:1463–1470.
162. Buellesfeld L, Gerckens U, Grube E. Percutaneous implantation of the first repositionable aortic valve prosthesis in a patient with severe aortic stenosis. *Catheter Cardiovasc Interv* 2008;**71**:579–584.
163. Schofer J, Schluter M, Treede H, Franzen OW, Tübler T, Pascotto A, Low RI, Bolling SF, Meinertz T, Reichenspurner H. Retrograde transarterial implantation of a nonmetallic aortic valve prosthesis in high-surgical-risk patients with severe aortic stenosis: a first-in-man feasibility and safety study. *Circ Cardiovasc Interv* 2008;**1**:126–133.
164. Nietlispach F, Wijesinghe N, Wood D, Carere RG, Webb JG. Current balloon-expandable transcatheter heart valve and delivery systems. *Catheter Cardiovasc Interv* 2010;**75**:295–300.
165. Falk V, Schwammenthal EE, Kempfert J, Linke A, Schuler G, Mohr FW, Walther T. New anatomically oriented transapical aortic valve implantation. *Ann Thorac Surg* 2009;**87**:925–926.
166. Feldman T, Wasserman HS, Herrmann HC, Gray W, Block PC, Whitlow P, St Goar F, Rodriguez L, Silvestry F, Schwartz A, Sanborn TA, Conrado JA, Foster E. Percutaneous mitral valve repair using the edge-to-edge technique: six-month results of the EVEREST Phase 1 Clinical trial. *J Am Coll Cardiol* 2005;**46**:2134–2140.
167. Feldman T, Kar S, Rinaldi M, Fail P, Hermiller J, Smalling R, Whitlow PL, Gray W, Low R, Herrmann HC, Lim S, Foster E, Glower D; EVEREST Investigators. Percutaneous mitral repair with the MitraClip system: safety and mid-term durability in the initial EVEREST (Endovascular Valve Edge-to-Edge REpair Study) cohort. *J Am Coll Cardiol* 2009;**54**:686–694.
168. Feldman T, Mauri L, Foster E, Glower D, on behalf of the EVEREST II Investigators. Endovascular valve edge-to-edge repair study (EVEREST II) randomized clinical trial: primary safety and efficacy endpoints. ACC late breaking clinical trial (March 14, 2010).